



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB122384

Analyst By : JIGNESH
Supervisor Review By : Iwona
Slope : 99.1
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/07/2022	08:25
2	CAL2	1	Water	NA	NA	20.7	7.00	10/07/2022	08:26
3	CAL3	1	Water	NA	NA	20.7	10.03	10/07/2022	08:27
4	ICV	1	Water	NA	NA	20.7	6.98	10/07/2022	08:30
5	CCV1	1	Water	NA	NA	20.6	2.01	10/07/2022	08:32
6	N5043-02	1	Water	NA	NA	20.2	7.26	10/07/2022	08:35
7	N5043-02DUP	1	Water	NA	NA	20.3	7.27	10/07/2022	08:36
8	N5043-03	1	Water	NA	NA	20.4	6.87	10/07/2022	08:39
9	N5045-01	1	Water	NA	NA	21.7	5.79	10/07/2022	08:42
10	CCV2	1	Water	NA	NA	20.8	12.03	10/07/2022	08:45

WORKLIST(Hardcopy Internal Chain)

VB122384

WorkList Name : CORROSIONITY N5043 W WorkList ID : 163878 Department : Wet-Chemistry Date : 10-07-2022 08:13:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/06/2022	Water	N5043-02	Corrosivity	Cool 4 deg C	PSEG03	C11	1112	10/06/2022	9040C
10/06/2022	Water	N5043-03	Corrosivity	Cool 4 deg C	PSEG03	C11	COMP-1	10/06/2022	9040C
10/06/2022	Water	N5045-01	Corrosivity	Cool 4 deg C	PSEG03	C11	8249	10/06/2022	9040C

Date/Time 10/07/22 08:15
 Raw Sample Received by: JO (WC)
 Raw Sample Relinquished by: OP SR

Date/Time 10/07/22 10:00
 Raw Sample Received by: OP SR
 Raw Sample Relinquished by: JO (WC)