



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122397

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/07/2022	16:13
2	CAL2	1	Water	NA	NA	20.7	7.00	10/07/2022	16:14
3	CAL3	1	Water	NA	NA	20.7	10.03	10/07/2022	16:15
4	ICV	1	Water	NA	NA	20.7	6.98	10/07/2022	15:16
5	CCV1	1	Water	NA	NA	20.7	2.01	10/07/2022	16:20
6	N5039-01	1	Water	NA	NA	20.9	6.12	10/07/2022	16:22
7	N5051-01	1	Water	NA	NA	22.2	6.78	10/07/2022	16:24
8	N5052-01	1	Water	NA	NA	20.1	7.73	10/07/2022	16:30
9	N5052-02	1	Water	NA	NA	20.2	7.98	10/07/2022	16:32
10	N5054-01	1	Water	NA	NA	20.1	8.30	10/07/2022	16:35
11	N5054-02	1	Water	NA	NA	20.2	8.11	10/07/2022	16:36
12	N5062-01	1	Water	NA	NA	20.1	7.04	10/07/2022	16:38
13	N5062-01DUP	1	Water	NA	NA	20.8	7.05	10/07/2022	16:39
14	CCV2	1	Water	NA	NA	20.8	12.03	10/07/2022	16:40