

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: JIGNESHRun Number: LB122399Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP98799

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB122399BL		1	21.3	0.85	0.91	10/07/2022	16:00
2	LB122399BS	1413	1	21.2	1327.00	1430.85	10/07/2022	16:05
3	N5039-01		1	21.6	0.98	1.05	10/07/2022	16:10
4	N5039-01DUP		1	21.5	0.97	1.04	10/07/2022	16:15

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

LB122399

WorkList Name : CONDUCTANCE-100722 WorkList ID : 163898 Department : Wet-Chemistry Date : 10-07-2022 09:32:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/20/2022	Water	N5039-01	Conductance	Cool 4 deg C	STER17	N11	JS9600	10/06/2022	SM2510 B

Date/Time 10/7/22 15:40
 Raw Sample Received by: 12/21
 Raw Sample Relinquished by: 12/21

Date/Time 10/7/22 16:30
 Raw Sample Received by: 12/21
 Raw Sample Relinquished by: 12/21