

OPERATOR: JASWAL

Linear

 μ Abs.:

35871

A= 0.0000e+000

B= 2.7645e-004

C= -4.4582e-002

Rho= 0.9995522

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	$\sqrt{2}$
0.0	0.000	0.020	0.020	234	0.000	234					
0.2	0.200	0.213	0.013	930	0.0 %	930					7
2.5	2.500	2.429	-0.071	8946	0.0 %	8946					-3
5.0	5.000	4.947	-0.053	18057	0.0 %	18057					-1
7.5	7.500	7.720	0.220	28086	0.0 %	28086					3
10.0	10.000	9.872	-0.128	35871	0.0 %	35871					-1

LB122411 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin JASWAL

Date of Analysis: 08 Oct 2022 18:46:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	234	-	0.0000	7470A	PPB	08 Oct 2022 18:47:08	S	Std
0.2 - 1	50.2	930	-	0.2000	7470A	PPB	08 Oct 2022 18:49:25	S	Std
2.5 - 1	52.5	8946	-	2.5000	7470A	PPB	08 Oct 2022 18:51:42	S	Std
5.0 - 1	55.0	18057	-	5.0000	7470A	PPB	08 Oct 2022 18:53:59	S	Std
7.5 - 1	57.5	28086	-	7.5000	7470A	PPB	08 Oct 2022 18:56:15	S	Std
10.0 - 1	50.0	35871	-	10.0000	7470A	PPB	08 Oct 2022 18:58:32	S	Std
ICV66 - 1	ICV66	13530	3.6958	-	7470A	PPB	08 Oct 2022 19:05:35	U	SMPL
ICB66 - 1	ICB66	120	-0.0114	-	7470A	PPB	08 Oct 2022 19:07:50	U	SMPL
CCV47 - 1	CCV47	17812	4.8795	-	7470A	PPB	08 Oct 2022 19:10:06	U	SMPL
CCB47 - 1	CCB47	101	-0.0167	-	7470A	PPB	08 Oct 2022 19:12:21	U	SMPL
CRA - 1	CRA	901	0.2045	-	7470A	PPB	08 Oct 2022 19:14:37	U	SMPL
HighStd - 1	HighStd	34652	9.5349	-	7470A	PPB	08 Oct 2022 19:16:52	U	SMPL
ChkStd - 1	ChkStd	25670	7.0518	-	7470A	PPB	08 Oct 2022 19:19:08	U	SMPL
PB148204BL - 1	PBW	70	-0.0252	-	7470A	PPB	08 Oct 2022 19:21:26	U	SMPL
PB148204BS - 1	LCSW	16571	4.5364	-	7470A	PPB	08 Oct 2022 19:23:44	U	SMPL
N5008-01 - 1	COMP-1	-2	-0.0451	-	7470A	PPB	08 Oct 2022 19:26:00	U	SMPL
N5008-02 - 1	COMP-2	37	-0.0344	-	7470A	PPB	08 Oct 2022 19:28:18	U	SMPL
N5008-03 - 1	2011	49	-0.0310	-	7470A	PPB	08 Oct 2022 19:30:35	U	SMPL
N5008-04 - 1	2007	540	0.1047	-	7470A	PPB	08 Oct 2022 19:32:52	U	SMPL
N5008-05 - 1	COMP-3	-20	-0.0501	-	7470A	PPB	08 Oct 2022 19:35:07	U	SMPL
CCV48 - 1	CCV48	19793	5.4272	-	7470A	PPB	08 Oct 2022 19:37:23	U	SMPL
CCB48 - 1	CCB48	118	-0.0120	-	7470A	PPB	08 Oct 2022 19:39:38	U	SMPL
N5014-16 - 1	COMP-1	86	-0.0208	-	7470A	PPB	08 Oct 2022 19:41:54	U	SMPL
N5014-17 - 1	COMP-2	81	-0.0222	-	7470A	PPB	08 Oct 2022 19:44:10	U	SMPL
N5019-01 - 1	SASP2-T-3-COMP	85	-0.0211	-	7470A	PPB	08 Oct 2022 19:46:26	U	SMPL
N5019-01DUP - 1	SASP2-T-3-COMPDUP	75	-0.0238	-	7470A	PPB	08 Oct 2022 19:48:41	U	SMPL
N5019-01MS - 1	SASP2-T-3-COMPMS	16248	4.4471	-	7470A	PPB	08 Oct 2022 19:50:57	U	SMPL
N5019-01MSD - 1	SASP2-T-3-COMPMSD	16399	4.4889	-	7470A	PPB	08 Oct 2022 19:53:13	U	SMPL
N5027-01 - 1	9-X-22-USED-Z-GREEN	1658	0.4138	-	7470A	PPB	08 Oct 2022 19:55:30	U	SMPL
N5027-02 - 1	9-X-23-USED-Z-GREEN	1225	0.2941	-	7470A	PPB	08 Oct 2022 19:57:48	U	SMPL
N5027-03 - 1	9-X-24-USED-Z-GREEN	955	0.2194	-	7470A	PPB	08 Oct 2022 20:00:05	U	SMPL
N5027-04 - 1	9-X-25-USED-Z-GREEN	730	0.1572	-	7470A	PPB	08 Oct 2022 20:02:21	U	SMPL
CCV49 - 1	CCV49	18877	5.1739	-	7470A	PPB	08 Oct 2022 20:04:36	U	SMPL
CCB49 - 1	CCB49	92	-0.0191	-	7470A	PPB	08 Oct 2022 20:06:52	U	SMPL
N5027-05 - 1	9-X-26-USED-Z-GREEN	1141	0.2708	-	7470A	PPB	08 Oct 2022 20:09:08	U	SMPL
N5027-06 - 1	9-X-27-USED-Z-GREEN	28	-0.0368	-	7470A	PPB	08 Oct 2022 20:11:24	U	SMPL
N5027-07 - 1	9-X-28-USED-Z-GREEN	95	-0.0183	-	7470A	PPB	08 Oct 2022 20:13:39	U	SMPL
N5027-08 - 1	9-X-29-USED-Z-GREEN	119	-0.0117	-	7470A	PPB	08 Oct 2022 20:15:55	U	SMPL
N5027-09 - 1	9-X-30-USED-Z-GREEN	126	-0.0097	-	7470A	PPB	08 Oct 2022 20:18:11	U	SMPL
PB148206BL - 1	PBW	219	0.0160	-	7470A	PPB	08 Oct 2022 20:20:28	U	SMPL
PB148206BS - 1	LCSW	16436	4.4991	-	7470A	PPB	08 Oct 2022 20:22:45	U	SMPL
N5006-02 - 1	COMP	471	0.0856	-	7470A	PPB	08 Oct 2022 20:25:02	U	SMPL
N5031-01 - 1	5869-3	16	-0.0402	-	7470A	PPB	08 Oct 2022 20:27:19	U	SMPL
N5031-02 - 1	5869-4	4	-0.0435	-	7470A	PPB	08 Oct 2022 20:29:35	U	SMPL
CCV50 - 1	CCV50	18828	5.1604	-	7470A	PPB	08 Oct 2022 20:36:17	U	SMPL
CCB50 - 1	CCB50	142	-0.0053	-	7470A	PPB	08 Oct 2022 20:38:32	U	SMPL
N5033-01 - 1	596C-1	-33	-0.0537	-	7470A	PPB	08 Oct 2022 20:40:48	U	SMPL
N5033-02 - 1	596C-2	-29	-0.0526	-	7470A	PPB	08 Oct 2022 20:43:04	U	SMPL
N5033-02DUP - 1	596C-2DUP	-28	-0.0523	-	7470A	PPB	08 Oct 2022 20:45:20	U	SMPL
N5033-02MS - 1	596C-2MS	15017	4.1068	-	7470A	PPB	08 Oct 2022 21:08:59	U	SMPL
N5033-02MSD - 1	596C-2MSD	15700	4.2957	-	7470A	PPB	08 Oct 2022 21:11:15	U	SMPL
CCV51 - 1	CCB51	18014	4.9354	-	7470A	PPB	08 Oct 2022 21:13:31	U	SMPL
CCB51 - 1	CCB51	96	-0.0180	-	7470A	PPB	08 Oct 2022 21:15:48	U	SMPL
N5042-04 - 1	EOR-0617	315	0.0425	-	7470A	PPB	08 Oct 2022 21:18:06	U	SMPL
N5042-06 - 1	EOR-0627	-60	-0.0612	-	7470A	PPB	08 Oct 2022 21:20:23	U	SMPL
N5042-07 - 1	EOR-0632	116	-0.0125	-	7470A	PPB	08 Oct 2022 21:22:39	U	SMPL
N5042-08 - 1	EOR-0631	93	-0.0189	-	7470A	PPB	08 Oct 2022 21:29:30	U	SMPL
N5051-01 - 1	WW-MANHOLE-1007	15	-0.0404	-	7470A	PPB	08 Oct 2022 21:31:46	U	SMPL
CCV52 - 1	CCV52	18071	4.9511	-	7470A	PPB	08 Oct 2022 21:34:02	U	SMPL
CCB52 - 1	CCB52	115	-0.0128	-	7470A	PPB	08 Oct 2022 21:36:17	U	SMPL
N5052-01 - 1	551B-1	-16	-0.0490	-	7470A	PPB	08 Oct 2022 21:38:33	U	SMPL
N5052-02 - 1	551B-2	6	-0.0429	-	7470A	PPB	08 Oct 2022 21:40:49	U	SMPL
N5054-01 - 1	510A-1	2	-0.0440	-	7470A	PPB	08 Oct 2022 21:43:05	U	SMPL
N5054-02 - 1	510A-2	-5	-0.0460	-	7470A	PPB	08 Oct 2022 21:45:21	U	SMPL
N5062-01 - 1	KS-MANHOLE-1007	-19	-0.0498	-	7470A	PPB	08 Oct 2022 21:47:38	U	SMPL
PB148207BL - 1	PBW	195	0.0093	-	7470A	PPB	08 Oct 2022 21:49:56	U	SMPL
PB148207BS - 1	LCSW	17177	4.7040	-	7470A	PPB	08 Oct 2022 21:52:12	U	SMPL
N5043-01 - 1	1108	-12	-0.0479	-	7470A	PPB	08 Oct 2022 21:54:28	U	SMPL
N5043-02 - 1	1112	-90	-0.0695	-	7470A	PPB	08 Oct 2022 21:56:45	U	SMPL
N5043-03 - 1	COMP-1	-95	-0.0708	-	7470A	PPB	08 Oct 2022 21:59:01	U	SMPL
CCV53 - 1	CCV53	18648	5.1106	-	7470A	PPB	08 Oct 2022 22:04:36	U	SMPL
CCB53 - 1	CCB53	112	-0.0136	-	7470A	PPB	08 Oct 2022 22:06:52	U	SMPL
N5045-01 - 1	8249	24	-0.0379	-	7470A	PPB	08 Oct 2022 22:09:09	U	SMPL
N5045-01DUP - 1	8249DUP	47	-0.0316	-	7470A	PPB	08 Oct 2022 22:11:25	U	SMPL
N5045-01MS - 1	8249MS	16048	4.3919	-	7470A	PPB	08 Oct 2022 22:16:54	U	SMPL
N5045-01MSD - 1	8249MSD	14205	3.8824	-	7470A	PPB	08 Oct 2022 22:19:10	U	SMPL
N5019-01LX5 - 1		59	-0.0283	-	7470A	PPB	08 Oct 2022 22:21:27	U	SMPL
N5019-01A - 1		15575	4.2611	-	7470A	PPB	08 Oct 2022 22:23:45	U	SMPL
N5045-01LX5 - 1		62	-0.0274	-	7470A	PPB	08 Oct 2022 22:26:01	U	SMPL
N5041-01A - 1		15608	4.2702	-	7470A	PPB	08 Oct 2022 22:28:18	U	SMPL
N5033-02LX5 - 1		128	-0.0092	-	7470A	PPB	08 Oct 2022 22:30:34	U	SMPL

LB122411 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Oct 2022 18:46:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
N5033-02A - 1		15653	4.2827		-7470A	PPB	08 Oct 2022 22:32:51	U	SMPL
CCV54 - 1	CCV54	18443	5.0539		-7470A	PPB	08 Oct 2022 22:35:08	U	SMPL
CCB54 - 1	CCB54	107	-0.0150		-7470A	PPB	08 Oct 2022 22:37:27	U	SMPL
PB148171TB - 1	LEB171	30	-0.0363		-7470A	PPB	08 Oct 2022 22:39:45	U	SMPL
PB148164TB - 1	LEB164	41	-0.0332		-7470A	PPB	08 Oct 2022 22:42:01	U	SMPL
CCV55 - 1	CCV55	18606	5.0990		-7470A	PPB	08 Oct 2022 22:47:58	U	SMPL
CCB55 - 1	CCB55	132	-0.0081		-7470A	PPB	08 Oct 2022 22:50:14	U	SMPL