



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122420

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/11/2022	08:25
2	CAL2	1	Water	NA	NA	20.7	7.00	10/11/2022	08:26
3	CAL3	1	Water	NA	NA	20.7	10.03	10/11/2022	08:27
4	ICV	1	Water	NA	NA	20.7	6.98	10/11/2022	08:30
5	CCV1	1	Water	NA	NA	20.7	2.01	10/11/2022	08:32
6	N5075-01	1	Water	NA	NA	20.4	5.23	10/11/2022	08:35
7	N5075-01DUP	1	Water	NA	NA	20.5	5.24	10/11/2022	08:36
8	N5075-02	1	Water	NA	NA	20.3	7.30	10/11/2022	08:39
9	N5087-04	1	Water	NA	NA	20.4	6.82	10/11/2022	08:40
10	N5087-05	1	Water	NA	NA	20.1	8.03	10/11/2022	08:42
11	N5087-06	1	Water	NA	NA	20.1	7.31	10/11/2022	08:45
12	CCV2	1	Water	NA	NA	20.7	12.03	10/11/2022	08:49