


SFAM01.1 LB122425, INSTRUMENT ID: CV1. ANALYSIS DATE: 10/11/2022
 OPERATOR: JASWAL



A= 0.0000e+000
 B= 2.5607e-004 *slope*
 C= 2.0955e-002 *y-intercept*
 Rho= 0.9997307
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>y.D</i>	<i>%D</i>
0.00	0.000	0.049	0.049	110	0.000	110						
0.05	0.050					185						
0.20	0.200	0.199	-0.001	697	0.0 %	697	0					
2.50	2.500	2.362	-0.138	9141	0.0 %	9141						
5.00	5.000	5.142	0.142	20000	0.0 %	20000						
7.50	7.500	7.454	-0.046	29026	0.0 %	29026						
10.0	10.000	9.994	-0.006	38946	0.0 %	38946						

JB
10/20/22

LB122425, INSTRUMENT ID: CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.00 - 1	110 -			0 SFAM01.1	PPB	10/11/2022 12:49	Std
0.20 - 1	697 -			0.2 SFAM01.1	PPB	10/11/2022 12:52	Std
2.50 - 1	9141 -			2.5 SFAM01.1	PPB	10/11/2022 12:54	Std
5.00 - 1	20000 -			5 SFAM01.1	PPB	10/11/2022 12:56	Std
7.50 - 1	29026 -			7.5 SFAM01.1	PPB	10/11/2022 12:58	Std
10.0 - 1	38946 -			10 SFAM01.1	PPB	10/11/2022 13:01	Std
ICV68 - 1 ICV68	14449	3.7209 -		SFAM01.1	PPB	10/11/2022 13:04	SMPL
ICB68 - 1 ICB68	-3	0.0202 -		SFAM01.1	PPB	10/11/2022 13:07	SMPL
CCV001 - 1 CCV001	20720	5.3267 -		SFAM01.1	PPB	10/11/2022 13:09	SMPL
CCB001 - 1 CCB001	-28	0.0138 -		SFAM01.1	PPB	10/11/2022 13:11	SMPL
PB148247E PBS247	270	0.0901 -		SFAM01.1	PPB	10/11/2022 13:14	SMPL
N4926-01 - MH0AC3	13650	3.5163 -		SFAM01.1	PPB	10/11/2022 13:16	SMPL
N4926-02 - MH0AC3D	11524	2.9719 -		SFAM01.1	PPB	10/11/2022 13:18	SMPL
N4926-03 - MH0AC3S	28818	7.4003 -		SFAM01.1	PPB	10/11/2022 13:20	SMPL
N4926-04 - MH0AC4	3847	1.0061 -		SFAM01.1	PPB	10/11/2022 13:23	SMPL
N4926-05 - MH0AC5	53087	13.6149 -		SFAM01.1	PPB	10/11/2022 13:25	SMPL
N4926-06 - MH0AC6	28149	7.229 -		SFAM01.1	PPB	10/11/2022 13:27	SMPL
N4926-07 - MH0AC7	21820	5.6084 -		SFAM01.1	PPB	10/11/2022 13:30	SMPL
N4926-08 - MH0AC8	32807	8.4218 -		SFAM01.1	PPB	10/11/2022 13:32	SMPL
N4926-09 - MH0AC9	27090	6.9579 -		SFAM01.1	PPB	10/11/2022 13:34	SMPL
N4926-10 - MH0AD0	35302	9.0607 -		SFAM01.1	PPB	10/11/2022 13:36	SMPL
N4926-11 - MH0AD1	3620	0.9479 -		SFAM01.1	PPB	10/11/2022 13:39	SMPL
N4926-12 - MH0AD2	5058	1.3162 -		SFAM01.1	PPB	10/11/2022 13:41	SMPL
N4926-13 - MH0AD3	5028	1.3085 -		SFAM01.1	PPB	10/11/2022 13:43	SMPL
N4926-14 - MH0AD4	2037	0.5426 -		SFAM01.1	PPB	10/11/2022 13:45	SMPL
N4926-15 - MH0AD5	67690	17.3543 -		SFAM01.1	PPB	10/11/2022 13:48	SMPL
N4926-16 - MH0AD6	26004	6.6798 -		SFAM01.1	PPB	10/11/2022 13:50	SMPL
N4926-17 - MH0AD7	213185	54.611 -		SFAM01.1	PPB	10/11/2022 13:52	SMPL
N4926-18 - MH0AD8	312686	80.0901 -		SFAM01.1	PPB	10/11/2022 13:55	SMPL
N4926-19 - MH0AD9	73098	18.7391 -		SFAM01.1	PPB	10/11/2022 13:57	SMPL
CCV002 - 1 CCV002	18359	4.7221 -		SFAM01.1	PPB	10/11/2022 14:00	SMPL
CCB002 - 1 CCB002	-71	0.0028 -		SFAM01.1	PPB	10/11/2022 14:02	SMPL
N4926-20 - MH0AE0	17363	4.4671 -		SFAM01.1	PPB	10/11/2022 14:04	SMPL
N4926-21 - MH0AE1	32058	8.23 -		SFAM01.1	PPB	10/11/2022 14:06	SMPL
N4926-22 - MH0AE2	75182	19.2727 -		SFAM01.1	PPB	10/11/2022 14:09	SMPL
PB148248E PBS248	-10	0.0184 -		SFAM01.1	PPB	10/11/2022 14:11	SMPL
N4927-01 - MH0AE3	1030	0.2847 -		SFAM01.1	PPB	10/11/2022 14:13	SMPL
N4927-02 - MH0AE4	20270	5.2115 -		SFAM01.1	PPB	10/11/2022 14:16	SMPL
N4927-03 - MH0AE4D	21709	5.58 -		SFAM01.1	PPB	10/11/2022 14:18	SMPL
N4926-05C MH0AC5	25685	6.5981 -		SFAM01.1	PPB	10/11/2022 14:22	SMPL
N4926-15C MH0AD5	13190	3.3985 -		SFAM01.1	PPB	10/11/2022 14:25	SMPL
N4926-17C MH0AD7	22601	5.8084 -		SFAM01.1	PPB	10/11/2022 14:27	SMPL
N4926-18C MH0AD8	13361	3.4423 -		SFAM01.1	PPB	10/11/2022 14:29	SMPL
N4926-19C MH0AD9	14965	3.853 -		SFAM01.1	PPB	10/11/2022 14:32	SMPL
N4926-22C MH0AE2	15163	3.9037 -		SFAM01.1	PPB	10/11/2022 14:34	SMPL
N4927-04 - MH0AE4S	33051	8.4843 -		SFAM01.1	PPB	10/11/2022 14:36	SMPL

LB122425, INSTRUMENT ID: CV1

CCV003 - 1 CCV003	19340	4.9733 -	SFAM01.1 PPB	10/11/2022 14:38 SMPL
CCB003 - 1 CCB003	-35	0.012 -	SFAM01.1 PPB	10/11/2022 14:41 SMPL