



PERCENT SOLID

Supervisor: Sohil  
Analyst: jignesh  
Date: 10/28/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:20  
In Date: 10/11/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 10/12/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB122426

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| N5015-01 | C1BE4           | 1      | 0.91           | 9.04         | 9.95                    | 6.39                      | 60.6    |          |
| N5015-02 | C1BE1           | 2      | 0.93           | 9.04         | 9.97                    | 8.31                      | 81.6    |          |
| N5015-03 | C1BE1MS         | 3      | 0.93           | 9.04         | 9.97                    | 8.31                      | 81.6    |          |
| N5015-04 | C1BE1MSD        | 4      | 0.93           | 9.04         | 9.97                    | 8.31                      | 81.6    |          |
| N5015-05 | C1BE5           | 5      | 0.92           | 9.02         | 9.94                    | 7.17                      | 69.3    |          |
| N5015-06 | C1BE6           | 6      | 0.91           | 8.77         | 9.68                    | 7.1                       | 70.6    |          |
| N5015-07 | C1BG4           | 7      | 0.95           | 8.96         | 9.91                    | 7.05                      | 68.1    |          |
| N5015-08 | C1BG0           | 8      | 0.93           | 9.03         | 9.96                    | 7.12                      | 68.5    |          |
| N5015-09 | C1BF8           | 9      | 0.91           | 9.07         | 9.98                    | 7.36                      | 71.1    |          |
| N5015-10 | C1BF9           | 10     | 0.91           | 9.05         | 9.96                    | 6.77                      | 64.8    |          |
| N5015-11 | C1BE7           | 11     | 0.91           | 9.05         | 9.96                    | 5.47                      | 50.4    |          |
| N5015-12 | C1BE8           | 12     | 0.91           | 8.95         | 9.86                    | 5.18                      | 47.7    |          |
| N5015-13 | C1BE9           | 13     | 0.91           | 9.03         | 9.94                    | 6.5                       | 61.9    |          |
| N5015-14 | C1BF0           | 14     | 0.95           | 9.01         | 9.96                    | 5.46                      | 50.1    |          |
| N5015-15 | C1BF3           | 15     | 0.93           | 9.01         | 9.94                    | 7.83                      | 76.6    |          |
| N5015-16 | C1BG1           | 16     | 0.91           | 9.04         | 9.95                    | 4.00                      | 34.2    |          |
| N5015-17 | C1BG2           | 17     | 0.91           | 9.06         | 9.97                    | 4.02                      | 34.3    |          |
| N5015-18 | C1BF2           | 18     | 0.91           | 9.06         | 9.97                    | 7.96                      | 77.8    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$