



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122429

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	10/11/2022	12:10
2	N5075-01		1	12	100.00	>212.0	>212.0	10/11/2022	12:30
3	N5075-02		1	14	100.00	>212.0	>212.0	10/11/2022	13:00
4	N5087-04		1	13	100.00	>212.0	>212.0	10/11/2022	13:30
5	N5087-05		1	14	100.00	>212.0	>212.0	10/11/2022	14:00
6	N5087-06		1	15	100.00	>212.0	>212.0	10/11/2022	14:30
7	N5093-03		1	14	100.00	>212.0	>212.0	10/11/2022	15:00
8	N5093-04		1	12	100.00	>212.0	>212.0	10/11/2022	15:30
9	N5095-01		1	13	100.00	>212.0	>212.0	10/11/2022	16:00
10	N5096-05		1	14	100.00	>212.0	>212.0	10/11/2022	16:30
11	N5096-06		1	13	100.00	>212.0	>212.0	10/11/2022	17:00
12	N5096-07		1	13	100.00	>212.0	>212.0	10/11/2022	17:30
13	N5096-07DUP		1	12	100.00	>212.0	>212.0	10/11/2022	18:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB122429

WorkList Name : Hex-1011

WorkList ID : 163977

Department : Wet-Chemistry

Date : 10-11-2022 12:05:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/17/2022	Water	N5075-01	D Flash Point	Cool 4 deg C	PSEG03	C11	CLI-0637	10/10/2022	1010B
10/17/2022	Water	N5075-02	D Flash Point	Cool 4 deg C	PSEG03	C11	M21015	10/10/2022	1010B
10/17/2022	Water	N5087-04	B Flash Point	Cool 4 deg C	PSEG03	C11	ORA-1512	10/10/2022	1010B
10/17/2022	Water	N5087-05	D Flash Point	Cool 4 deg C	PSEG03	C11	31857	10/10/2022	1010B
10/17/2022	Water	N5087-06	D Flash Point	Cool 4 deg C	PSEG03	C11	ORA-1509	10/10/2022	1010B
10/17/2022	Water	N5093-03	F Flash Point	Cool 4 deg C	PSEG03	C11	D2735	10/10/2022	1010B
10/17/2022	Water	N5093-04	F Flash Point	Cool 4 deg C	PSEG03	C11	COMP(D2724-D2723-D2722)	10/10/2022	1010B
10/17/2022	Water	N5095-01	D Flash Point	Cool 4 deg C	PSEG03	C11	FRAC-TANK-FM786	10/10/2022	1010B
10/17/2022	Water	N5096-05	F Flash Point	Cool 4 deg C	PSEG03	C11	COMP(2692-2691)	10/10/2022	1010B
10/17/2022	Water	N5096-06	F Flash Point	Cool 4 deg C	PSEG03	C11	2673	10/10/2022	1010B
10/17/2022	Water	N5096-07	F Flash Point	Cool 4 deg C	PSEG03	C11	2693	10/10/2022	1010B

Date/Time 10/11/22 12:05
 Raw Sample Received by: SJ (wc)
 Raw Sample Relinquished by: CP (SM)

Date/Time 10/11/22 18:30
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: SJ (wc)