


SFAM01.1 LB122434, INSTRUMENT ID:CV1, DATE OF ANALYSIS:10/11/2022

OPERATOR: JASWAL



A= 0.0000e+000
 B= 2.4106e-004 *slope*
 C= 2.6189e-002 *y-intercept*
 Rho= 0.9992526
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%LD
0.00	0.000	0.035	0.035	35	0.000	35					
0.05	0.050					185	0				
0.20	0.200	0.208	0.008	753	0.0 %	753					
2.50	2.500	2.346	-0.154	9623	0.0 %	9623					
5.00	5.000	5.258	0.258	21702	0.0 %	21702					
7.50	7.500	7.327	-0.173	30286	0.0 %	30286					
10.0	10.000	10.027	0.027	41487	0.0 %	41487					

Handwritten notes:
 100
 4
 6
 5
 2
 0

Sample ID	Extended I μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.00 - 1		35 -		0 SFAM01.1	PPB	10/11/2022 15:57	Std
0.20 - 1		753 -		0.2 SFAM01.1	PPB	10/11/2022 15:59	Std
2.50 - 1		9623 -		2.5 SFAM01.1	PPB	10/11/2022 16:02	Std
5.00 - 1		21702 -		5 SFAM01.1	PPB	10/11/2022 16:04	Std
7.50 - 1		30286 -		7.5 SFAM01.1	PPB	10/11/2022 16:06	Std
10.0 - 1		41487 -		10 SFAM01.1	PPB	10/11/2022 16:08	Std
ICV70 - 1	ICV70	15122	3.6715 -	SFAM01.1	PPB	10/11/2022 16:14	SMPL
ICB70 - 1	ICB70	67	0.0423 -	SFAM01.1	PPB	10/11/2022 16:26	SMPL
CCV004 - 1	CCV004	20893	5.0627 -	SFAM01.1	PPB	10/11/2022 16:28	SMPL
CCB004 - 1	CCB004	-73	0.0086 -	SFAM01.1	PPB	10/11/2022 16:30	SMPL
PB148245E	PBW	230	0.0816 -	SFAM01.1	PPB	10/11/2022 16:32	SMPL
N4828-01 -	MCB8P3	497	0.146 -	SFAM01.1	PPB	10/11/2022 16:35	SMPL
N4828-02 -	MCB8P3D	270	0.0913 -	SFAM01.1	PPB	10/11/2022 16:37	SMPL
N4828-03 -	MCB8P3S	4915	1.211 -	SFAM01.1	PPB	10/11/2022 16:39	SMPL
N4828-04 -	MCB8P5	137	0.0592 -	SFAM01.1	PPB	10/11/2022 16:42	SMPL
N4828-05 -	MCB8P7	175	0.0684 -	SFAM01.1	PPB	10/11/2022 16:44	SMPL
N4828-06 -	MCB8Q5	225	0.0804 -	SFAM01.1	PPB	10/11/2022 16:46	SMPL
N4828-07 -	MCB8Q9	165	0.066 -	SFAM01.1	PPB	10/11/2022 16:48	SMPL
N4828-08 -	MCB8R1	181	0.0698 -	SFAM01.1	PPB	10/11/2022 16:51	SMPL
N4828-09 -	MCB8R4	162	0.0652 -	SFAM01.1	PPB	10/11/2022 16:53	SMPL
N4828-10 -	MCB8S5	171	0.0674 -	SFAM01.1	PPB	10/11/2022 16:55	SMPL
N4828-11 -	MCB8T4	268	0.0908 -	SFAM01.1	PPB	10/11/2022 16:57	SMPL
N4828-12 -	MCB8T5	236	0.0831 -	SFAM01.1	PPB	10/11/2022 17:00	SMPL
N5057-01 -	BGS25	162	0.0652 -	SFAM01.1	PPB	10/11/2022 17:02	SMPL
N5057-02 -	BGS25D	174	0.0681 -	SFAM01.1	PPB	10/11/2022 17:04	SMPL
PB148246E	PBW	209	0.0766 -	SFAM01.1	PPB	10/11/2022 17:09	SMPL
N4829-01 -	MCB8P4	205	0.0756 -	SFAM01.1	PPB	10/11/2022 17:11	SMPL
N4829-02 -	MCB8P4D	179	0.0693 -	SFAM01.1	PPB	10/11/2022 17:13	SMPL
N4829-03 -	MCB8P4S	4730	1.1664 -	SFAM01.1	PPB	10/11/2022 17:16	SMPL
CCV005 - 1	CCV005	20329	4.9267 -	SFAM01.1	PPB	10/11/2022 17:18	SMPL
CCB005 - 1	CCB005	-39	0.0168 -	SFAM01.1	PPB	10/11/2022 17:20	SMPL
N4829-04 -	MCB8P8	155	0.0636 -	SFAM01.1	PPB	10/11/2022 17:23	SMPL
N4829-05 -	MCB8P9	156	0.0638 -	SFAM01.1	PPB	10/11/2022 17:25	SMPL
N4829-06 -	MCB8Q6	204	0.0754 -	SFAM01.1	PPB	10/11/2022 17:27	SMPL
N4829-07 -	MCB8R0	147	0.0616 -	SFAM01.1	PPB	10/11/2022 17:29	SMPL
N4829-08 -	MCB8R2	144	0.0609 -	SFAM01.1	PPB	10/11/2022 17:32	SMPL
N4829-09 -	MCB8T6	217	0.0785 -	SFAM01.1	PPB	10/11/2022 17:34	SMPL
N4829-10 -	MCB8T7	197	0.0737 -	SFAM01.1	PPB	10/11/2022 17:36	SMPL
N4953-01 -	MCB8T8	205	0.0756 -	SFAM01.1	PPB	10/11/2022 17:38	SMPL
N4954-01 -	MCB8W1	198	0.0739 -	SFAM01.1	PPB	10/11/2022 17:41	SMPL
N4994-01 -	BGS11	152	0.0628 -	SFAM01.1	PPB	10/11/2022 17:43	SMPL
N4994-02 -	BGS14	158	0.0643 -	SFAM01.1	PPB	10/11/2022 17:45	SMPL
N4994-03 -	BGS14D	158	0.0643 -	SFAM01.1	PPB	10/11/2022 17:48	SMPL
N4994-04 -	BGS14S	4762	1.1741 -	SFAM01.1	PPB	10/11/2022 17:50	SMPL
N5056-01 -	BGS24	137	0.0592 -	SFAM01.1	PPB	10/11/2022 17:52	SMPL
N5056-02 -	BGS24D	151	0.0626 -	SFAM01.1	PPB	10/11/2022 17:54	SMPL

N5056-03 - BGS24S	4851	1.1956 -	SFAM01.1 PPB	10/11/2022 17:57 SMPL
N5057-03 - BGS25S	4863	1.1985 -	SFAM01.1 PPB	10/11/2022 17:59 SMPL
CCV006 - 1 CCV006	20903	5.0651 -	SFAM01.1 PPB	10/11/2022 18:01 SMPL
CCB006 - 1 CCB006	-58	0.0122 -	SFAM01.1 PPB	10/11/2022 18:03 SMPL