


245.1 LB122436, INSTRUMENT ID:CV1, ANALYSIS DATE:10/11/2022

OPERATOR: JASWAL



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	-31	0.000	-31				4.0
0.05	0.050	0.036	-0.014	185	0.0 %	185				-28
0.2	0.200	0.162	-0.038	649	0.0 %	649				-16
2.5	2.500	2.366	-0.134	8804	0.0 %	8804				-5
5.0	5.000	5.406	0.406	20051	0.0 %	20051				8
7.5	7.500	7.467	-0.033	27673	0.0 %	27673				-14
10.0	10.000	9.835	-0.165	36432	0.0 %	36432				-2

LB122436 INSTRUMENT ID: CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Oct 2022 18:42:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-31	-	0.0000	245.1	PPB	11 Oct 2022 19:37:02	S	Std
0.05 - 1	50.05	185	-	0.0500	245.1	PPB	11 Oct 2022 19:39:19	S	Std
0.2 - 1	50.2	649	-	0.2000	245.1	PPB	11 Oct 2022 19:41:37	S	Std
2.5 - 1	52.5	8804	-	2.5000	245.1	PPB	11 Oct 2022 19:43:54	S	Std
5.0 - 1	55.0	20051	-	5.0000	245.1	PPB	11 Oct 2022 19:46:10	S	Std
7.5 - 1	57.5	27673	-	7.5000	245.1	PPB	11 Oct 2022 19:48:26	S	Std
10.0 - 1	510.0	36432	-	10.0000	245.1	PPB	11 Oct 2022 19:50:44	S	Std
ICV71 - 1	ICV71	15457	4.1646	-245.1		PPB	11 Oct 2022 20:05:26	U	SMPL
ICB71 - 1	ICB71	-78	-0.0348	-245.1		PPB	11 Oct 2022 20:07:41	U	SMPL
CCV10 - 1	CCV10	20185	5.4427	-245.1		PPB	11 Oct 2022 20:09:57	U	SMPL
CCB10 - 1	CCB10	-91	-0.0383	-245.1		PPB	11 Oct 2022 20:12:12	U	SMPL
CRA - 1	CRA	238	0.0506	-245.1		PPB	11 Oct 2022 20:14:29	U	SMPL
HighStd - 1	HighStd	39390	10.6342	-245.1		PPB	11 Oct 2022 20:16:44	U	SMPL
ChkStd - 1	ChkStd	27293	7.3641	-245.1		PPB	11 Oct 2022 20:19:00	U	SMPL
PB148283BL - 1	PBW	-120	-0.0461	-245.1		PPB	11 Oct 2022 20:21:17	U	SMPL
PB148283BS - 1	LCSW	16231	4.3738	-245.1		PPB	11 Oct 2022 20:23:35	U	SMPL
N4989-01 - 1	402	-53	-0.0280	-245.1		PPB	11 Oct 2022 20:25:51	U	SMPL
N4989-01DUP - 1	402DUP	8	-0.0115	-245.1		PPB	11 Oct 2022 20:28:09	U	SMPL
N4989-01MS - 1	402MS	16308	4.3947	-245.1		PPB	11 Oct 2022 20:30:25	U	SMPL
N4989-01MSD - 1	402MSD	15658	4.2190	-245.1		PPB	11 Oct 2022 20:32:42	U	SMPL
N4989-01LX5 - 1	402L	-75	-0.0340	-245.1		PPB	11 Oct 2022 20:34:58	U	SMPL
N4989-01A-1	CCV11	18593	5.0123	-245.1		PPB	11 Oct 2022 20:37:14	U	SMPL
GCV11 - 1	CCB11	-51	-0.0275	-245.1		PPB	11 Oct 2022 20:39:29	U	SMPL
CCB11 - 1	GCV11	15139	4.0787	-245.1		PPB	11 Oct 2022 20:41:45	U	SMPL
CCV12 - 1	CCV12	18856	5.0834	-245.1		PPB	11 Oct 2022 20:44:01	U	SMPL
CCB12 - 1	CCB12	-53	-0.0280	-245.1		PPB	11 Oct 2022 20:46:18	U	SMPL

9
10/11/22