



PERCENT SOLID

Supervisor: Sohil  
Analyst: JIGNESH  
Date: 10/13/2022

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:10  
In Date: 10/12/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:15  
Out Date: 10/13/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB122442

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| N5070-01 | C0BK4           | 1      | 0.91           | 8.88         | 9.79                    | 7.45                      | 73.6    |          |
| N5070-02 | C0BK5           | 2      | 0.95           | 9.01         | 9.96                    | 8.4                       | 82.7    |          |
| N5070-03 | C0BK7           | 3      | 0.91           | 8.88         | 9.79                    | 7.54                      | 74.7    |          |
| N5070-04 | C0BL1           | 4      | 0.91           | 8.77         | 9.68                    | 7.96                      | 80.4    |          |
| N5070-05 | C0BL7           | 5      | 0.98           | 8.57         | 9.55                    | 7.35                      | 74.3    |          |
| N5070-06 | C0BM2           | 6      | 0.94           | 8.98         | 9.92                    | 8.3                       | 82.0    |          |
| N5070-07 | C0BM5           | 7      | 0.91           | 8.93         | 9.84                    | 7.93                      | 78.6    |          |
| N5070-08 | C0BQ6           | 8      | 0.93           | 9.04         | 9.97                    | 8.17                      | 80.1    |          |
| N5070-09 | C0BQ7           | 9      | 0.93           | 9.04         | 9.97                    | 9.03                      | 89.6    |          |
| N5070-10 | C0BQ9           | 10     | 0.93           | 8.85         | 9.78                    | 8.2                       | 82.1    |          |
| N5070-11 | C0BR4           | 11     | 0.95           | 8.67         | 9.62                    | 8.19                      | 83.5    |          |
| N5070-12 | C0BS2           | 12     | 0.96           | 8.68         | 9.64                    | 7.95                      | 80.5    |          |
| N5070-13 | C0BS5           | 13     | 0.97           | 8.92         | 9.89                    | 7.98                      | 78.6    |          |
| N5070-17 | VHBLK002        | 14     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | VHBLK    |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$