



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122453

Analyst By : JIGNESH

Supervisor Review By : Sohil

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/13/2022	08:15
2	CAL2	1	Water	NA	NA	20.7	7.00	10/13/2022	08:16
3	CAL3	1	Water	NA	NA	20.7	10.03	10/13/2022	08:16
4	ICV	1	Water	NA	NA	20.7	6.98	10/13/2022	08:20
5	CCV1	1	Water	NA	NA	20.8	2.01	10/13/2022	08:22
6	N5114-05	1	Water	NA	NA	20.6	7.47	10/13/2022	08:24
7	N5116-03	1	Water	NA	NA	21.6	9.23	10/13/2022	08:30
8	N5116-04	1	Water	NA	NA	21.8	9.17	10/13/2022	08:32
9	N5117-03	1	Water	NA	NA	21.8	7.62	10/13/2022	08:39
10	N5127-01	1	Water	NA	NA	21.1	5.29	10/13/2022	08:40
11	N5127-03	1	Water	NA	NA	21.7	5.29	10/13/2022	08:42
12	N5127-05	1	Water	NA	NA	21.7	5.23	10/13/2022	08:45
13	N5128-01	1	Water	NA	NA	20.4	7.39	10/13/2022	08:46
14	N5128-02	1	Water	NA	NA	20.6	7.55	10/13/2022	08:50
15	N5130-01	1	Water	NA	NA	20.6	8.04	10/13/2022	08:53
16	CCV2	1	Water	NA	NA	20.7	12.03	10/13/2022	08:55
17	N5130-01DUP	1	Water	NA	NA	20.8	8.05	10/13/2022	08:56
18	CCV3	1	Water	NA	NA	20.8	2.01	10/13/2022	09:05

VB 122453

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n5130 WorkList ID : 164016 Department : Wet-Chemistry Date : 10-13-2022 08:08:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/12/2022	Water	N5114-05	pH	Cool 4 deg C	PSEG03	C11	FMC-0104	10/12/2022	9040C
10/12/2022	Water	N5116-03	pH	Cool 4 deg C	PSEG03	C11	SWRN-3	10/12/2022	9040C
10/12/2022	Water	N5116-04	pH	Cool 4 deg C	PSEG03	C11	SWRN-4	10/12/2022	9040C
10/12/2022	Water	N5117-03	pH	Cool 4 deg C	PSEG03	C11	34219	10/12/2022	9040C
10/12/2022	Water	N5127-01	pH	Cool 4 deg C	PSEG03	C11	34247	10/12/2022	9040C
10/12/2022	Water	N5127-03	pH	Cool 4 deg C	PSEG03	C11	34248	10/12/2022	9040C
10/12/2022	Water	N5127-05	pH	Cool 4 deg C	PSEG03	C11	34249	10/12/2022	9040C
10/11/2022	Water	N5128-01	pH	Cool 4 deg C	ENVI70	C11	265751-3	10/11/2022	9040C
10/11/2022	Water	N5128-02	pH	Cool 4 deg C	ENVI70	C11	265751-4	10/11/2022	9040C
10/11/2022	Water	N5130-01	pH	Cool 4 deg C	ENVI70	C11	557-E-1	10/11/2022	9040C

Date/Time 10-13-22 08:10
 Raw Sample Received by: 70/60C1
 Raw Sample Relinquished by: CR SR

Date/Time 10-13-22 10:00
 Raw Sample Received by: AP SR
 Raw Sample Relinquished by: 70/60C1