



TOTAL SOLIDS - SM2540B

SUPERVISOR: Iwona

ANALYST: JIGNESH

Date: 10/13/2022

Run Number: LB122461

BalanceID: WC SC-6

OvenID: WC OVEN-1

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 10/13/2022 11:00 TEMP1 OUT: 103 °C 10/13/2022 12:00
TEMP2 IN: 103 °C 10/13/2022 12:30 TEMP2 OUT: 104 °C 10/13/2022 13:30
TEMP3 IN: 104 °C 10/13/2022 16:25 TEMP3 OUT: 103 °C 10/14/2022 08:00
TEMP4 IN: 104 °C 10/14/2022 08:30 TEMP4 OUT: 103 °C 10/14/2022 10:00

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB122461BL	LB122461BL	96.8783	96.8783	100	96.8783	96.8783	96.8783	0.0000	0
2	N5133-03	HAS-WATER-1013	97.6789	97.6789	100	97.7696	97.7696	97.7696	0.0907	907
3	N5133-03DUP	HAS-WATER-1013DUP	90.6584	90.6584	100	90.7489	90.7489	90.7489	0.0905	905
4	N5141-01	SV34211L	88.8757	88.8757	100	89.4698	89.4698	89.4698	0.5941	5941

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB 122461

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS WN5133 WorkList ID : 164042 Department : Wet-Chemistry Date : 10-13-2022 14:14:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/20/2022	Water	N5133-03	TS <u>R</u>	Cool 4 deg C	PSEG03	C11	HAS-WATER-1013	10/13/2022	SM2540 B
10/18/2022	Water	N5141-01	TS <u>C</u>	Cool 4 deg C	TULL02	D11	SV34211L	10/13/2022	SM2540 B

Date/Time 10.13.22 14:20
Raw Sample Received by: 706017
Raw Sample Relinquished by: h.c. sam

Date/Time 10.13.22 16:30
Raw Sample Received by: h.c. sam
Raw Sample Relinquished by: 7P.11111