



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 10/13/2022 16:25 TEMP1 OUT: 103 °C 10/14/2022 08:00
TEMP2 IN: 104 °C 10/14/2022 08:30 TEMP2 OUT: 103 °C 10/14/2022 10:00
TEMP3 IN: 540 °C 10/14/2022 10:30 TEMP3 OUT: 540 °C 10/14/2022 12:00
TEMP4 IN: 560 °C 10/14/2022 12:30 TEMP4 OUT: 560 °C 10/14/2022 14:00

Run Number: LB122462

SUPERVISOR: Iwona

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB122462BL	96.8783	96.8783	100	96.8783	96.8783	96.8783	96.8783	0.0000	0
2	N5133-03	97.6789	97.6789	100	97.7696	97.7696	97.7088	97.7088	0.0608	608
3	N5133-03DUP	90.6584	90.6584	100	90.7489	90.7489	90.6857	90.6857	0.0632	632

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105 °C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105 °C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

VB 122462

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TVS N5133

WorkList ID : 164043

Department : Wet-Chemistry

Date : 10-13-2022 14:15:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/20/2022	Water	N5133-03	TVS	✓	Cool 4 deg C	PSEG03 C11	HAS-WATER-1013	10/13/2022	SM2540 G

Date/Time 10-13-22 14:20
Raw Sample Received by: 701051
Raw Sample Relinquished by: Z.C. 6m

Date/Time 10-13-22 16:30
Raw Sample Received by: Z-C 6m
Raw Sample Relinquished by: 701051