


7470A LB122463, INSTRUMENT ID:CV1, ANALYSIS DATE:10/13/2022

OPERATOR: JASWAL



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.089	-0.089	160	0.000	160				
0.2	0.200	0.178	-0.022	1086	0.0 %	1086				
2.5	2.500	2.436	-0.064	8912	0.0 %	8912				
5.0	5.000	5.366	0.366	19069	0.0 %	19069				
7.5	7.500	7.468	-0.032	26355	0.0 %	26355				
10.0	10.000	9.840	-0.160	34578	0.0 %	34578				

1.2
-11
-3
7
-0.4
-2

LB122463, INSTRUMENT IS: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Oct 2022 15:35:03

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	160	-	0.000	7470A	PPB	13 Oct 2022 16:01:20	S	Std
0.2 - 1	50.2	1086	-	0.200	7470A	PPB	13 Oct 2022 16:03:38	S	Std
2.5 - 1	52.5	8912	-	2.500	7470A	PPB	13 Oct 2022 16:05:55	S	Std
5.0 - 1	55.0	19069	-	5.000	7470A	PPB	13 Oct 2022 16:08:12	S	Std
7.5 - 1	57.5	26355	-	7.500	7470A	PPB	13 Oct 2022 16:10:28	S	Std
10.0 - 1	59.0	34578	-	10.000	7470A	PPB	13 Oct 2022 16:12:45	S	Std
ICV76 - 1	ICV76	13199	3.6728	-	7470A	PPB	13 Oct 2022 16:19:27	U	SMPL
ICB76 - 1	ICB76	534	0.0192	-	7470A	PPB	13 Oct 2022 16:21:42	U	SMPL
CCV54 - 1	CCV54	18354	5.1599	-	7470A	PPB	13 Oct 2022 16:23:56	U	SMPL
CCB54 - 1	CCB54	576	0.0313	-	7470A	PPB	13 Oct 2022 16:30:06	U	SMPL
CRA - 1	CRA	1078	0.1761	-	7470A	PPB	13 Oct 2022 16:49:16	U	SMPL
HighStd - 1	HighStd	33408	9.5026	-	7470A	PPB	13 Oct 2022 16:51:31	U	SMPL
ChkStd - 1	ChkStd	23642	6.6853	-	7470A	PPB	13 Oct 2022 16:53:46	U	SMPL
CCV55 - 1	CCV55	17679	4.9651	-	7470A	PPB	13 Oct 2022 16:57:43	U	SMPL
CCB55 - 1	CCB55	408	-0.0171	-	7470A	PPB	13 Oct 2022 16:59:59	U	SMPL
PB148219BL - 1	PBW	311	-0.0451	-	7470A	PPB	13 Oct 2022 17:02:16	U	SMPL
PB148219BS - 1	LCSW	14628	4.0850	-	7470A	PPB	13 Oct 2022 17:04:33	U	SMPL
N4892-02 - 1	TOPSOIL-STORMWATER-DENALI	421	-0.0134	-	7470A	PPB	13 Oct 2022 17:06:50	U	SMPL
N4892-02DUP - 1	TOPSOIL-STORMWATER-DENALIDUP	298	-0.0489	-	7470A	PPB	13 Oct 2022 17:09:08	U	SMPL
N4892-02MS - 1	TOPSOIL-STORMWATER-DENALIMS	11840	3.2807	-	7470A	PPB	13 Oct 2022 17:11:23	U	SMPL
N4892-02MSD - 1	TOPSOIL-STORMWATER-DENALIMSD	12119	3.3612	-	7470A	PPB	13 Oct 2022 17:13:38	U	SMPL
PB148321BL - 1	PBW	179	-0.0832	-	7470A	PPB	13 Oct 2022 17:15:55	U	SMPL
CCV56 - 1	CCV56	19041	5.3580	-	7470A	PPB	13 Oct 2022 17:24:05	U	SMPL
CCB56 - 1	CCB56	149	-0.0918	-	7470A	PPB	13 Oct 2022 17:26:20	U	SMPL
PB148321BS - 1	LCSW	14558	4.0648	-	7470A	PPB	13 Oct 2022 17:28:37	U	SMPL
N5075-01 - 1	CLI-0637	308	-0.0460	-	7470A	PPB	13 Oct 2022 17:30:53	U	SMPL
N5075-02 - 1	M21015	709	0.0697	-	7470A	PPB	13 Oct 2022 17:33:11	U	SMPL
N5087-05 - 1	31857	142	-0.0939	-	7470A	PPB	13 Oct 2022 17:35:28	U	SMPL
N5087-08 - 1	ORA-1509	10929	3.0179	-	7470A	PPB	13 Oct 2022 17:37:44	U	SMPL
N5095-01 - 1	FRAC-TANK-FM786	485	0.0051	-	7470A	PPB	13 Oct 2022 17:40:01	U	SMPL
N5095-01DUP - 1	FRAC-TANK-FM786DUP	622	0.0446	-	7470A	PPB	13 Oct 2022 17:42:20	U	SMPL
N5095-01MS - 1	FRAC-TANK-FM786MS	14142	3.9448	-	7470A	PPB	13 Oct 2022 17:44:36	U	SMPL
N5095-01MSD - 1	FRAC-TANK-FM786MSD	14976	4.1854	-	7470A	PPB	13 Oct 2022 17:49:43	U	SMPL
N5116-03 - 1	SWRN-3	68	-0.1152	-	7470A	PPB	13 Oct 2022 17:51:58	U	SMPL
CCV57 - 1	CCV57	16136	4.5200	-	7470A	PPB	13 Oct 2022 17:57:04	U	SMPL
CCB57 - 1	CCB57	209	-0.0745	-	7470A	PPB	13 Oct 2022 17:59:19	U	SMPL
N5116-04 - 1	SWRN-4	404	-0.0183	-	7470A	PPB	13 Oct 2022 18:01:37	U	SMPL
N5128-01 - 1	265751-3	597	0.0374	-	7470A	PPB	13 Oct 2022 18:03:52	U	SMPL
N5128-02 - 1	265751-4	572	0.0302	-	7470A	PPB	13 Oct 2022 18:06:08	U	SMPL
N5130-01 - 1	557-E-1	468	0.0002	-	7470A	PPB	13 Oct 2022 18:08:25	U	SMPL
N4892-02LX5 - 1	TOPSOIL-STORMWATER-DENALIL	249	-0.0630	-	7470A	PPB	13 Oct 2022 18:10:41	U	SMPL
N4892-02A - 1	TOPSOIL-STORMWATER-DENALIA	13611	3.7916	-	7470A	PPB	13 Oct 2022 18:12:58	U	SMPL
PB148219TB - 1	LEB219	188	-0.0806	-	7470A	PPB	13 Oct 2022 18:15:15	U	SMPL
N5095-01LX5 - 1	FRAC-TANK-FM786L	520	0.0152	-	7470A	PPB	13 Oct 2022 18:17:32	U	SMPL
N5095-01A - 1	FRAC-TANK-FM786A	13581	3.7830	-	7470A	PPB	13 Oct 2022 18:19:48	U	SMPL
CCV58 - 1	CCV58	17194	4.8252	-	7470A	PPB	13 Oct 2022 18:22:04	U	SMPL
CCB58 - 1	CCB58	360	-0.0310	-	7470A	PPB	13 Oct 2022 18:29:56	U	SMPL