



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB122466

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	10/13/2022	15:45
2	CAL2	1	Water	NA	NA	20.4	7.00	10/13/2022	15:46
3	CAL3	1	Water	NA	NA	20.7	10.03	10/13/2022	15:47
4	ICV	1	Water	NA	NA	20.8	6.99	10/13/2022	15:50
5	CCV1	1	Water	NA	NA	20.7	2.01	10/13/2022	16:00
6	N5133-02	1	Solid	20.03	20	22.6	8.38	10/13/2022	16:01
7	N5140-02	1	Solid	20.04	20	22.9	7.91	10/13/2022	16:03
8	N5140-04	1	Solid	20.02	20	22.8	11.26	10/13/2022	16:10
9	N5140-06	1	Solid	20.05	20	22.9	9.65	10/13/2022	16:12
10	N5140-08	1	Solid	20.03	20	22.7	8.06	10/13/2022	16:15
11	N5140-10	1	Solid	20.04	20	23.2	7.03	10/13/2022	16:20
12	N5140-12	1	Solid	20.03	20	22.8	8.41	10/13/2022	16:22
13	N5140-14	1	Solid	20.01	20	22.8	8.39	10/13/2022	16:25
14	N5140-14DUP	1	Solid	20.02	20	22.9	8.40	10/13/2022	16:26
15	CCV2	1	Water	NA	NA	20.7	12.03	10/13/2022	16:30

JP VP 122466 JP
VP 122466

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N5140 WorkList ID : 164047 Department : Wet-Chemistry Date : 10-13-2022 15:26:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/13/2022	Solid	N5133-02	Corrosivity	Cool 4 deg C	PSEG03	C11	HAS-SOIL-1013	10/13/2022	9045D
10/12/2022	Solid	N5140-02	Corrosivity	Cool 4 deg C	LIRO01	C11	SB-01-COMP	10/12/2022	9045D
10/12/2022	Solid	N5140-04	Corrosivity	Cool 4 deg C	LIRO01	C11	SB-02-COMP	10/12/2022	9045D
10/12/2022	Solid	N5140-06	Corrosivity	Cool 4 deg C	LIRO01	C11	SB-03-COMP	10/12/2022	9045D
10/12/2022	Solid	N5140-08	Corrosivity	Cool 4 deg C	LIRO01	C11	SB-04-COMP	10/12/2022	9045D
10/12/2022	Solid	N5140-10	Corrosivity	Cool 4 deg C	LIRO01	C11	SB-05-COMP	10/12/2022	9045D
10/12/2022	Solid	N5140-12	Corrosivity	Cool 4 deg C	LIRO01	C11	SB-06-COMP	10/12/2022	9045D
10/12/2022	Solid	N5140-14	Corrosivity	Cool 4 deg C	LIRO01	C11	SB-07-COMP	10/12/2022	9045D

Date/Time 10.13.22 15:35
Raw Sample Received by: JP/WW
Raw Sample Relinquished by: JD CSM

Date/Time 10.13.22 17:10
Raw Sample Received by: JD CSM
Raw Sample Relinquished by: JP/WW