



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122476

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	10/14/2022	08:35
2	N5110-02		1	13	100.00	>212.0	>212.0	10/14/2022	09:00
3	N5114-05		1	14	100.00	>212.0	>212.0	10/14/2022	09:30
4	N5116-03		1	12	100.00	>212.0	>212.0	10/14/2022	10:00
5	N5116-04		1	13	100.00	>212.0	>212.0	10/14/2022	10:30
6	N5117-03		1	14	100.00	>212.0	>212.0	10/14/2022	11:00
7	N5127-01		1	15	100.00	>212.0	>212.0	10/14/2022	11:30
8	N5127-03		1	13	100.00	>212.0	>212.0	10/14/2022	12:00
9	N5127-05		1	14	100.00	>212.0	>212.0	10/14/2022	12:30
10	N5133-03		1	14	100.00	>212.0	>212.0	10/14/2022	13:00
11	N5141-01		1	13	100.00	>212.0	>212.0	10/14/2022	13:30
12	N5141-01DUP		1	12	100.00	>212.0	>212.0	10/14/2022	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB122476

WorkList Name : FP-1014

WorkList ID : 164070

Department : Wet-Chemistry

Date : 10-14-2022 08:30:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/19/2022	Water	N5110-02	E Flash Point	Cool 4 deg C	PSEG03	C11	MANSFIELD-OILY-Water	10/12/2022	1010B
10/19/2022	Water	N5114-05	E Flash Point	Cool 4 deg C	PSEG03	C11	FMC-0104	10/12/2022	1010B
10/19/2022	Water	N5116-03	E Flash Point	Cool 4 deg C	PSEG03	C11	SWRN-3	10/12/2022	1010B
10/19/2022	Water	N5116-04	E Flash Point	Cool 4 deg C	PSEG03	C11	SWRN-4	10/12/2022	1010B
10/19/2022	Water	N5117-03	C Flash Point	Cool 4 deg C	PSEG03	C11	34219	10/12/2022	1010B
10/19/2022	Water	N5127-01	C Flash Point	Cool 4 deg C	PSEG03	C11	34247	10/12/2022	1010B
10/19/2022	Water	N5127-03	C Flash Point	Cool 4 deg C	PSEG03	C11	34248	10/12/2022	1010B
10/19/2022	Water	N5127-05	C Flash Point	Cool 4 deg C	PSEG03	C11	34249	10/12/2022	1010B
10/20/2022	Water	N5133-03	N Flash Point	Cool 4 deg C	PSEG03	C11	HAS-WATER-1013	10/13/2022	1010B
10/18/2022	Water	N5141-01	I Flash Point	Cool 4 deg C	TULL02	D11	SV34211L	10/13/2022	1010B

Date/Time 10/14/22 8:30
 Raw Sample Received by: sj(wc)
 Raw Sample Relinquished by: pp(ext)

Date/Time 10/14/22 14:30
 Raw Sample Received by: pp(ext)
 Raw Sample Relinquished by: sj(wc)