


SFAM01.1 LB122491, INSTRUMENT ID: CV1, DATE OF ANALYSIS:10/16/2022

OPERATOR: JASWAL



Linear

A= 0.0000e+000

B= 2.3210e-004 *slope*

C= -5.0086e-002 *y-intercept*

Rho= 0.9999922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% <i>15000</i>
0.00	0.000	0.002	0.002	223	0.000	223					
0.05	0.050					185					
0.20	0.200	0.210	0.010	1120	0.0 %	1120	0				
2.50	2.500	2.502	0.002	10995	0.0 %	10995					
5.00	5.000	4.985	-0.015	21695	0.0 %	21695					
7.50	7.500	7.479	-0.021	32437	0.0 %	32437					
10.0	10.000	10.023	0.023	43397	0.0 %	43397					

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.00 - 1	223	-		0 SFAM01.1	PPB	10/16/2022 18:34	Std
0.20 - 1	1120	-		0.2 SFAM01.1	PPB	10/16/2022 18:36	Std
2.50 - 1	10995	-		2.5 SFAM01.1	PPB	10/16/2022 18:38	Std
5.00 - 1	21695	-		5 SFAM01.1	PPB	10/16/2022 18:41	Std
7.50 - 1	32437	-		7.5 SFAM01.1	PPB	10/16/2022 18:43	Std
10.0 - 1	43397	-		10 SFAM01.1	PPB	10/16/2022 18:45	Std
ICV78 - 1 ICV78	17918	4.1087	-	SFAM01.1	PPB	10/16/2022 19:02	SMPL
ICB78 - 1 ICB78	161	-0.0127	-	SFAM01.1	PPB	10/16/2022 19:04	SMPL
CCV001 - 1 CCV001	21303	4.8944	-	SFAM01.1	PPB	10/16/2022 19:07	SMPL
CCB001 - 1 CCB001	142	-0.0171	-	SFAM01.1	PPB	10/16/2022 19:09	SMPL
PB148373E PBS373	201	-0.0034	-	SFAM01.1	PPB	10/16/2022 19:11	SMPL
N5063-01 - EYB71	663	0.1038	-	SFAM01.1	PPB	10/16/2022 19:13	SMPL
N5063-02 - EYBB1	604	0.0901	-	SFAM01.1	PPB	10/16/2022 19:16	SMPL
N5063-03 - EYBB2	545	0.0764	-	SFAM01.1	PPB	10/16/2022 19:18	SMPL
N5063-04 - EYBB7	704	0.1133	-	SFAM01.1	PPB	10/16/2022 19:20	SMPL
N5063-05 - EYBB8	540	0.0752	-	SFAM01.1	PPB	10/16/2022 19:22	SMPL
N5063-06 - EYBC0	832	0.143	-	SFAM01.1	PPB	10/16/2022 19:25	SMPL
N5063-07 - EYBD2	-81	-0.0689	-	SFAM01.1	PPB	10/16/2022 19:27	SMPL
N5063-08 - EYBD3	517	0.0699	-	SFAM01.1	PPB	10/16/2022 19:32	SMPL
N5063-09 - EYBD5	5022	1.1155	-	SFAM01.1	PPB	10/16/2022 19:35	SMPL
N5063-10 - EYBD6	464	0.0576	-	SFAM01.1	PPB	10/16/2022 19:37	SMPL
N5063-11 - EYBD8	531	0.0732	-	SFAM01.1	PPB	10/16/2022 19:39	SMPL
N5063-12 - EYBE0	1102	0.2057	-	SFAM01.1	PPB	10/16/2022 19:41	SMPL
N5063-13 - EYBE1	621	0.0941	-	SFAM01.1	PPB	10/16/2022 19:44	SMPL
N5063-14 - EYBE2	1543	0.308	-	SFAM01.1	PPB	10/16/2022 19:46	SMPL
N5063-15 - EYBE3	640	0.0985	-	SFAM01.1	PPB	10/16/2022 19:48	SMPL
N5063-16 - EYBE4	1404	0.2758	-	SFAM01.1	PPB	10/16/2022 19:50	SMPL
N5063-17 - EYB59	741	0.1219	-	SFAM01.1	PPB	10/16/2022 19:53	SMPL
CCV002 - 1 CCV002	21401	4.9172	-	SFAM01.1	PPB	10/16/2022 19:55	SMPL
CCB002 - 1 CCB002	96	-0.0278	-	SFAM01.1	PPB	10/16/2022 19:57	SMPL
N5063-18 - EYB69	1992	0.4123	-	SFAM01.1	PPB	10/16/2022 19:59	SMPL
N5063-19 - EYBC1	1178	0.2233	-	SFAM01.1	PPB	10/16/2022 20:02	SMPL
N5063-20 - MEYBD1	940	0.1681	-	SFAM01.1	PPB	10/16/2022 20:04	SMPL
N5063-21 - MEYBD1D	895	0.1576	-	SFAM01.1	PPB	10/16/2022 20:06	SMPL
N5063-22 - MEYBD1S	14009	3.2015	-	SFAM01.1	PPB	10/16/2022 20:12	SMPL
CCV003 - 1 CCV003	21281	4.8893	-	SFAM01.1	PPB	10/16/2022 20:14	SMPL
CCB003 - 1 CCB003	236	0.0047	-	SFAM01.1	PPB	10/16/2022 20:17	SMPL