



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122505

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/17/2022	15:05
2	CAL2	1	Water	NA	NA	20.7	7.00	10/17/2022	15:06
3	CAL3	1	Water	NA	NA	20.7	10.03	10/17/2022	15:09
4	ICV	1	Water	NA	NA	20.7	6.98	10/17/2022	15:12
5	CCV1	1	Water	NA	NA	20.7	2.01	10/17/2022	15:15
6	N5172-01	1	Solid	20.03	20	21.3	8.16	10/17/2022	15:20
7	N5172-01DUP	1	Solid	20.04	20	21.4	8.17	10/17/2022	15:21
8	CCV2	1	Water	NA	NA	20.8	12.03	10/17/2022	15:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N5172 WorkList ID : 164139 Department : Wet-Chemistry Date : 10-17-2022 14:58:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/17/2022	Solid	N5172-01	pH	Cool 4 deg C	PSEG05	D11	BU-02-101722	10/17/2022	9045D

Date/Time 10-17-22 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10-17-22 16:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]