

LB122511


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43114



A= 0.0000e+000

B= 2.3095e-004

C= -7.8115e-003

Rho= 0.9999380

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0,1,2
0.0	0.000	0.004	0.004	49	0.000	49					
0.2	0.200	0.206	0.006	924	0.0 %	924					
2.5	2.500	2.465	-0.035	10705	0.0 %	10705					
5.0	5.000	4.998	-0.002	21673	0.0 %	21673					
7.5	7.500	7.579	0.079	32851	0.0 %	32851					
10.0	10.000	9.950	-0.050	43114	0.0 %	43114					

LB122511 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Oct 2022 09:21:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	49	-	0.0000	7470A	PPB	18 Oct 2022 09:42:42	S	Std
0.2 - 1	50.2	924	-	0.2000	7470A	PPB	18 Oct 2022 09:44:59	S	Std
2.5 - 1	52.5	10705	-	2.5000	7470A	PPB	18 Oct 2022 09:47:16	S	Std
5.0 - 1	55	21673	-	5.0000	7470A	PPB	18 Oct 2022 09:49:33	S	Std
7.5 - 1	57.5	32851	-	7.5000	7470A	PPB	18 Oct 2022 09:51:49	S	Std
10.0 - 1	59	43114	-	10.0000	7470A	PPB	18 Oct 2022 09:54:05	S	Std
ICV81 - 1	ICV81	16970	3.9115	-	7470A	PPB	18 Oct 2022 09:56:54	U	SMPL
ICB81 - 1	ICB81	-42	-0.0175	-	7470A	PPB	18 Oct 2022 09:59:09	U	SMPL
CCV35 - 1	CCV35	22224	5.1249	-	7470A	PPB	18 Oct 2022 10:01:25	U	SMPL
CCB35 - 1	CCB35	-40	-0.0170	-	7470A	PPB	18 Oct 2022 10:03:41	U	SMPL
CRA - 1	CRA	913	0.2030	-	7470A	PPB	18 Oct 2022 10:05:56	U	SMPL
HighStd - 1	HighStd	43571	10.0550	-	7470A	PPB	18 Oct 2022 10:08:11	U	SMPL
ChkStd - 1	ChkStd	30714	7.0857	-	7470A	PPB	18 Oct 2022 10:10:27	U	SMPL
PB148396BL - 1	PBW	-90	-0.0286	-	7470A	PPB	18 Oct 2022 10:12:45	U	SMPL
PB148396BS - 1	LCSW	17203	3.9653	-	7470A	PPB	18 Oct 2022 10:15:02	U	SMPL
N5110-02 - 1	MANSFIELD-OILY-Water	-43	-0.0177	-	7470A	PPB	18 Oct 2022 10:17:18	U	SMPL
N5110-02DUP - 1	MANSFIELD-OILY-WaterDUP	49	0.0035	-	7470A	PPB	18 Oct 2022 10:19:35	U	SMPL
N5110-02MS - 1	MANSFIELD-OILY-WaterMS	20224	4.6630	-	7470A	PPB	18 Oct 2022 10:21:52	U	SMPL
N5110-02MSD - 1	MANSFIELD-OILY-WaterMSD	19007	4.3819	-	7470A	PPB	18 Oct 2022 10:24:09	U	SMPL
N5159-01 - 1	SB-32	-11	-0.0104	-	7470A	PPB	18 Oct 2022 10:26:25	U	SMPL
CCV36 - 1	CCV36	22241	5.1288	-	7470A	PPB	18 Oct 2022 10:28:42	U	SMPL
CCB36 - 1	CCB36	-47	-0.0187	-	7470A	PPB	18 Oct 2022 10:30:57	U	SMPL
N5159-02 - 1	SB-33	57	0.0054	-	7470A	PPB	18 Oct 2022 10:33:14	U	SMPL
N5159-03 - 1	SB-34	75	0.0095	-	7470A	PPB	18 Oct 2022 10:35:29	U	SMPL
N5159-04 - 1	SB-35	113	0.0183	-	7470A	PPB	18 Oct 2022 10:37:44	U	SMPL
N5159-05 - 1	SB-36	75	0.0095	-	7470A	PPB	18 Oct 2022 10:40:00	U	SMPL
N5159-06 - 1	SB-37	53	0.0044	-	7470A	PPB	18 Oct 2022 10:42:16	U	SMPL
N5159-07 - 1	SB-38	84	0.0116	-	7470A	PPB	18 Oct 2022 10:44:32	U	SMPL
N5159-08 - 1	SB-39	55	0.0049	-	7470A	PPB	18 Oct 2022 10:46:48	U	SMPL
N5159-09 - 1	SB-40	66	0.0074	-	7470A	PPB	18 Oct 2022 10:49:05	U	SMPL
N5159-10 - 1	SB-41	85	0.0118	-	7470A	PPB	18 Oct 2022 10:51:23	U	SMPL
N5159-11 - 1	SB-42	55	0.0049	-	7470A	PPB	18 Oct 2022 10:53:38	U	SMPL
CCV37 - 1	CCV37	21743	5.0138	-	7470A	PPB	18 Oct 2022 10:55:53	U	SMPL
CCB37 - 1	CCB37	-58	-0.0212	-	7470A	PPB	18 Oct 2022 10:58:09	U	SMPL
N5159-12 - 1	SB-43	54	0.0047	-	7470A	PPB	18 Oct 2022 11:00:26	U	SMPL
N5159-13 - 1	SB-44	69	0.0081	-	7470A	PPB	18 Oct 2022 11:02:42	U	SMPL
N5159-14 - 1	SB-45	63	0.0067	-	7470A	PPB	18 Oct 2022 11:04:58	U	SMPL
N5159-14DUP - 1	SB-45DUP	63	0.0067	-	7470A	PPB	18 Oct 2022 11:07:14	U	SMPL
N5159-14MS - 1	SB-45MS	18817	4.3380	-	7470A	PPB	18 Oct 2022 11:09:30	U	SMPL
N5159-14MSD - 1	SB-45MSD	17387	4.0078	-	7470A	PPB	18 Oct 2022 11:11:46	U	SMPL
PB148397BL - 1	PBW	-37	-0.0164	-	7470A	PPB	18 Oct 2022 11:14:03	U	SMPL
PB148397BS - 1	LCSW	15609	3.5971	-	7470A	PPB	18 Oct 2022 11:16:19	U	SMPL
N5111-03 - 1	WC-3	-28	-0.0143	-	7470A	PPB	18 Oct 2022 11:18:35	U	SMPL
N5133-02 - 1	HAS-SOIL-1013	7	-0.0062	-	7470A	PPB	18 Oct 2022 11:20:52	U	SMPL
CCV38 - 1	CCV38	20933	4.8267	-	7470A	PPB	18 Oct 2022 11:23:09	U	SMPL
CCB38 - 1	CCB38	-89	-0.0284	-	7470A	PPB	18 Oct 2022 11:25:25	U	SMPL
N5140-02 - 1	SB-01-COMP	26	-0.0018	-	7470A	PPB	18 Oct 2022 11:27:41	U	SMPL
N5140-04 - 1	SB-02-COMP	18	-0.0037	-	7470A	PPB	18 Oct 2022 11:29:57	U	SMPL
N5140-06 - 1	SB-03-COMP	9	-0.0057	-	7470A	PPB	18 Oct 2022 11:32:14	U	SMPL
N5140-08 - 1	SB-04-COMP	36	0.0005	-	7470A	PPB	18 Oct 2022 11:34:29	U	SMPL
N5140-10 - 1	SB-05-COMP	90	0.0130	-	7470A	PPB	18 Oct 2022 11:36:45	U	SMPL
N5140-12 - 1	SB-06-COMP	26	-0.0018	-	7470A	PPB	18 Oct 2022 11:39:01	U	SMPL
N5140-14 - 1	SB-07-COMP	25	-0.0020	-	7470A	PPB	18 Oct 2022 11:41:17	U	SMPL
N5142-03 - 1	WC-1	32	-0.0004	-	7470A	PPB	18 Oct 2022 11:43:33	U	SMPL
N5142-06 - 1	WC-11	58	0.0056	-	7470A	PPB	18 Oct 2022 11:45:49	U	SMPL
N5142-09 - 1	WC-10	60	0.0060	-	7470A	PPB	18 Oct 2022 11:48:06	U	SMPL
CCV39 - 1	CCV39	21390	4.9323	-	7470A	PPB	18 Oct 2022 11:50:23	U	SMPL
CCB39 - 1	CCB39	-99	-0.0307	-	7470A	PPB	18 Oct 2022 11:52:38	U	SMPL
N5142-12 - 1	WC-9	47	0.0030	-	7470A	PPB	18 Oct 2022 11:54:54	U	SMPL
N5142-15 - 1	WC-8	121	0.0201	-	7470A	PPB	18 Oct 2022 11:57:10	U	SMPL
N5142-15DUP - 1	WC-8DUP	73	0.0090	-	7470A	PPB	18 Oct 2022 11:59:27	U	SMPL
N5142-15MS - 1	WC-8MS	18522	4.2699	-	7470A	PPB	18 Oct 2022 12:01:43	U	SMPL
N5142-15MSD - 1	WC-8MSD	16782	3.8680	-	7470A	PPB	18 Oct 2022 12:03:58	U	SMPL
PB148398BL - 1	PBW	-2	-0.0083	-	7470A	PPB	18 Oct 2022 12:06:14	U	SMPL
PB148398BS - 1	LCSW	18164	4.1872	-	7470A	PPB	18 Oct 2022 12:08:29	U	SMPL
N5133-03 - 1	HAS-WATER-1013	-80	-0.0263	-	7470A	PPB	18 Oct 2022 12:10:46	U	SMPL
N5141-01 - 1	SV34211L	73	0.0090	-	7470A	PPB	18 Oct 2022 12:13:03	U	SMPL
N5141-01DUP - 1	SV34211LDUP	104	0.0162	-	7470A	PPB	18 Oct 2022 12:15:20	U	SMPL
CCV40 - 1	CCV40	21357	4.9247	-	7470A	PPB	18 Oct 2022 12:17:38	U	SMPL
CCB40 - 1	CCB40	-61	-0.0219	-	7470A	PPB	18 Oct 2022 12:19:54	U	SMPL
N5141-01MS - 1	SV34211LMS	14161	3.2627	-	7470A	PPB	18 Oct 2022 12:22:10	U	SMPL
N5141-01MSD - 1	SV34211LMSD	14594	3.3627	-	7470A	PPB	18 Oct 2022 12:24:29	U	SMPL
N5145-03 - 1	MW-6-101322	-109	-0.0330	-	7470A	PPB	18 Oct 2022 12:26:46	U	SMPL
N5145-04 - 1	MW-7-101322	871	0.1933	-	7470A	PPB	18 Oct 2022 12:29:02	U	SMPL
N5145-05 - 1	MW-8-101322	33	-0.0002	-	7470A	PPB	18 Oct 2022 12:31:19	U	SMPL
N5145-06 - 1	MW-9-101322	8	-0.0060	-	7470A	PPB	18 Oct 2022 12:33:36	U	SMPL
N5147-01 - 1	557-E-2	-11	-0.0104	-	7470A	PPB	18 Oct 2022 12:35:52	U	SMPL
N5151-01 - 1	256317-5	-13	-0.0108	-	7470A	PPB	18 Oct 2022 12:38:08	U	SMPL
N5151-02 - 1	256317-6	-3	-0.0085	-	7470A	PPB	18 Oct 2022 12:40:25	U	SMPL
N5164-01 - 1	588A-1	49	0.0035	-	7470A	PPB	18 Oct 2022 12:42:43	U	SMPL
CCV41 - 1	CCV41	21366	4.9267	-	7470A	PPB	18 Oct 2022 12:45:01	U	SMPL
							18 Oct 2022 12:47:18	U	SMPL
							18 Oct 2022 12:49:34	U	SMPL

LB122511 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Oct 2022 09:21:22

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB41 - 1	CCB41	-11	-0.0104		-7470A	PPB	18 Oct 2022 12:51:50	U	SMPI
N5164-02 - 1	588A-2	26	-0.0018		-7470A	PPB	18 Oct 2022 12:54:06	U	SMPI
N5166-01 - 1	5388-1	19	-0.0034		-7470A	PPB	18 Oct 2022 12:56:23	U	SMPI
N5166-02 - 1	5388-2	47	0.0030		-7470A	PPB	18 Oct 2022 12:58:39	U	SMPL
N5170-10 - 1	COMP-TRE-1662-1667	979	0.2183		-7470A	PPB	18 Oct 2022 13:00:55	U	SMPL
PB148365TB - 1	PB148365TB	16	-0.0041		-7470A	PPB	18 Oct 2022 13:03:11	U	SMPL
PB148323TB - 1	PB148323TB	59	0.0058		-7470A	PPB	18 Oct 2022 13:05:28	U	SMPL
PB148322TB - 1	PB148322TB	81	0.0109		-7470A	PPB	18 Oct 2022 13:07:45	U	SMPL
CCV42 - 1	CCV42	21218	4.8926		-7470A	PPB	18 Oct 2022 13:10:02	U	SMPL
CCB42 - 1	CCB42	-14	-0.0110		-7470A	PPB	18 Oct 2022 13:12:19	U	SMPL
N5110-02LX5 - 1		-37	-0.0164		-7470A	PPB	18 Oct 2022 13:14:37	U	SMPL
N5110-02A - 1		19591	4.5168		-7470A	PPB	18 Oct 2022 13:16:53	U	SMPL
N5159-14LX5 - 1		-97	-0.0302		-7470A	PPB	18 Oct 2022 13:19:09	U	SMPL
N5159-14A - 1		16882	3.8911		-7470A	PPB	18 Oct 2022 13:21:26	U	SMPL
N5142-15LX5 - 1		-133	-0.0385		-7470A	PPB	18 Oct 2022 13:23:42	U	SMPL
N5142-15A - 1		16872	3.8888		-7470A	PPB	18 Oct 2022 13:26:00	U	SMPL
N5141-01LX5 - 1		-88	-0.0281		-7470A	PPB	18 Oct 2022 13:28:16	U	SMPL
N5141-01A - 1		17113	3.9445		-7470A	PPB	18 Oct 2022 13:30:33	U	SMPL
CCV43 - 1	CCV43	21254	4.9009		-7470A	PPB	18 Oct 2022 13:32:49	U	SMPL
CCB43 - 1	CCB43	-68	-0.0235		-7470A	PPB	18 Oct 2022 13:35:06	U	SMPL