

LB 122516


7471B

INSTRUMENT 10: CV 1

Linear

 μ Abs.:

52688



A= 0.0000e+000

B= 1.9044e-004

C= 1.4948e-003

Rho= 0.9998909

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.013	0.013	60	0.000	60					
0.2	0.200	0.223	0.023	1165	0.0 %	1165					12
2.5	2.500	2.431	-0.069	12759	0.0 %	12759					-3
5.0	5.000	5.073	0.073	26630	0.0 %	26630					1
7.5	7.500	7.424	-0.076	38974	0.0 %	38974					-1
10.0	10.000	10.036	0.036	52688	0.0 %	52688					0

LB122516 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Oct 2022 14:13:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	60	-	0.0000	7471B	PPB	18 Oct 2022 14:21:07	S	Std
0.2 - 1	50.2	1165	-	0.2000	7471B	PPB	18 Oct 2022 14:23:27	S	Std
2.5 - 1	52.5	12759	-	2.5000	7471B	PPB	18 Oct 2022 14:25:44	S	Std
5.0 - 1	55	26630	-	5.0000	7471B	PPB	18 Oct 2022 14:28:01	S	Std
7.5 - 1	57.5	38974	-	7.5000	7471B	PPB	18 Oct 2022 14:30:18	S	Std
10.0 - 1	510	52688	-	10.0000	7471B	PPB	18 Oct 2022 14:32:36	S	Std
ICV82 - 1	ICV82	21954	4.1825	-	7471B	PPB	18 Oct 2022 14:35:41	U	SMPL
ICB82 - 1	ICB82	-71	-0.0120	-	7471B	PPB	18 Oct 2022 14:37:56	U	SMPL
CCV44 - 1	CCV44	26881	5.1208	-	7471B	PPB	18 Oct 2022 14:40:13	U	SMPL
CCB44 - 1	CCB44	-111	-0.0196	-	7471B	PPB	18 Oct 2022 14:42:28	U	SMPL
CRA - 1	CRA	1165	0.2234	-	7471B	PPB	18 Oct 2022 14:44:45	U	SMPL
HighStd - 1	HighStd	52806	10.0580	-	7471B	PPB	18 Oct 2022 14:47:00	U	SMPL
ChkStd - 1	ChkStd	36075	6.8717	-	7471B	PPB	18 Oct 2022 14:49:16	U	SMPL
PB148395BL - 1	PBS	-158	-0.0286	-	7471B	PPB	18 Oct 2022 14:51:37	U	SMPL
PB148395BS - 1	LCSS	20814	3.9654	-	7471B	PPB	18 Oct 2022 14:53:57	U	SMPL
N5133-01 - 1	HAS-SOIL-1013	2240	0.4281	-	7471B	PPB	18 Oct 2022 14:56:14	U	SMPL
N5135-01 - 1	OK-1-101322	11620	2.2144	-	7471B	PPB	18 Oct 2022 14:58:33	U	SMPL
N5142-01 - 1	WC-1	1399	0.2679	-	7471B	PPB	18 Oct 2022 15:00:51	U	SMPL
N5142-01DUP - 1	WC-1DUP	1606	0.3073	-	7471B	PPB	18 Oct 2022 15:03:10	U	SMPL
N5142-01MS - 1	WC-1MS	21713	4.1366	-	7471B	PPB	18 Oct 2022 15:05:28	U	SMPL
CCV45 - 1	CCV45	26813	5.1078	-	7471B	PPB	18 Oct 2022 15:07:43	U	SMPL
CCB45 - 1	CCB45	-74	-0.0126	-	7471B	PPB	18 Oct 2022 15:10:00	U	SMPL
N5142-01MSD - 1	WC-1MSD	21981	4.1876	-	7471B	PPB	18 Oct 2022 15:12:18	U	SMPL
N5142-04 - 1	WC-11	607	0.1171	-	7471B	PPB	18 Oct 2022 15:14:33	U	SMPL
N5142-07 - 1	WC-10	2317	0.4428	-	7471B	PPB	18 Oct 2022 15:16:52	U	SMPL
N5142-10 - 1	WC-9	1095	0.2100	-	7471B	PPB	18 Oct 2022 15:19:08	U	SMPL
N5142-13 - 1	WC-8	2169	0.4146	-	7471B	PPB	18 Oct 2022 15:21:23	U	SMPL
N5143-01 - 1	EO-01-101322	2510	0.4795	-	7471B	PPB	18 Oct 2022 15:23:40	U	SMPL
N5160-01 - 1	HD-1-101422	1666	0.3188	-	7471B	PPB	18 Oct 2022 15:25:56	U	SMPL
N5162-01 - 1	TR-4-101422	2001	0.3826	-	7471B	PPB	18 Oct 2022 15:28:13	U	SMPL
N5170-01 - 1	1639	258	0.0506	-	7471B	PPB	18 Oct 2022 15:30:30	U	SMPL
N5170-02 - 1	TRE-1594	466	0.0902	-	7471B	PPB	18 Oct 2022 15:32:46	U	SMPL
CCV46 - 1	CCV46	26914	5.1271	-	7471B	PPB	18 Oct 2022 15:35:01	U	SMPL
CCB46 - 1	CCB46	-62	-0.0103	-	7471B	PPB	18 Oct 2022 15:37:17	U	SMPL
N5171-01 - 1	NB-08-101722	1843	0.3525	-	7471B	PPB	18 Oct 2022 15:39:35	U	SMPL
N5172-01 - 1	BU-02-101722	2323	0.4439	-	7471B	PPB	18 Oct 2022 15:41:50	U	SMPL
N5142-01LX5 - 1		207	0.0409	-	7471B	PPB	18 Oct 2022 15:44:06	U	SMPL
N5142-01A - 1		26061	4.9646	-	7471B	PPB	18 Oct 2022 15:46:21	U	SMPL
CCV47 - 1	CCV47	26941	5.1322	-	7471B	PPB	18 Oct 2022 15:48:37	U	SMPL
CCB47 - 1	CCB47	-67	-0.0113	-	7471B	PPB	18 Oct 2022 15:50:55	U	SMPL