



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB122519
Thermometer ID:

Reviewed By: Jignesh
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 750.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	10/18/2022	12:30
2	N5158-01		1	12	100.00	>212.0	>212.0	10/18/2022	13:00
3	N5170-03		1	14	100.00	>212.0	>212.0	10/18/2022	13:30
4	N5170-10		1	12	100.00	>212.0	>212.0	10/18/2022	14:00
5	N5170-11		1	13	100.00	>212.0	>212.0	10/18/2022	14:30
6	N5181-01		1	12	100.00	>212.0	>212.0	10/18/2022	15:00
7	N5181-01DUP		1	14	100.00	>212.0	>212.0	10/18/2022	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB1225T9

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flash 1018 WorkList ID : 164155 Department : Wet-Chemistry Date : 10-18-2022 12:28:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/21/2022	Water	N5158-01	Flash Point	Cool 4 deg C	PSEG03	D11	COMP-TOTE	10/14/2022	1010B
10/19/2022	Water	N5170-03	Flash Point	Cool 4 deg C	PSEG03	D11	TRE-1661	10/14/2022	1010B
10/19/2022	Water	N5170-10	Flash Point	Cool 4 deg C	PSEG03	D11	COMP-TRE-1662-1667	10/14/2022	1010B
10/19/2022	Water	N5170-11	Flash Point	Cool 4 deg C	PSEG03	D11	1275	10/14/2022	1010B
10/21/2022	Water	N5181-01	Flash Point	Cool 4 deg C	TULL02	D11	PIT-14-EXCAVATION-WATER	10/18/2022	1010B

Date/Time 10/18/2022 12:30
 Raw Sample Received by: 51 (WC)
 Raw Sample Relinquished by: CP (SM)

Date/Time 10/18/22
 Raw Sample Received by: CP (SM)
 Raw Sample Relinquished by: 51 (WC)