

LB122528


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45024



A= 0.0000e+000

B= 2.2444e-004

C= -7.7888e-003

Rho= 0.9989982

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	0.016	0.016	106	0.000	106					
0.2	0.200	0.207	0.007	957	0.0 %	957					3
2.5	2.500	2.422	-0.078	10828	0.0 %	10828					-3
5.0	5.000	5.250	0.250	23424	0.0 %	23424					5
7.5	7.500	7.208	-0.292	32148	0.0 %	32148					-4
10.0	10.000	10.097	0.097	45024	0.0 %	45024					1

LB122528 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Oct 2022 13:42:09

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	106	-	0.0000	7470A	PPB	19 Oct 2022 13:48:50	S	Std
0.2 - 1	50.2	957	-	0.2000	7470A	PPB	19 Oct 2022 13:51:08	S	Std
2.5 - 1	52.5	10828	-	2.5000	7470A	PPB	19 Oct 2022 13:53:24	S	Std
5.0 - 1	55	23424	-	5.0000	7470A	PPB	19 Oct 2022 13:55:43	S	Std
7.5 - 1	57.5	32148	-	7.5000	7470A	PPB	19 Oct 2022 13:58:00	S	Std
10.0 - 1	59	45024	-	10.0000	7470A	PPB	19 Oct 2022 14:08:03	S	Std
ICV83 - 1	ICV83	16279	3.6459	-	7470A	PPB	19 Oct 2022 14:10:47	U	SMPL
ICB83 - 1	ICB83	86	0.0115	-	7470A	PPB	19 Oct 2022 14:13:02	U	SMPL
CCV48 - 1	CCV48	20955	4.6954	-	7470A	PPB	19 Oct 2022 14:15:18	U	SMPL
CCB48 - 1	CCB48	77	0.0095	-	7470A	PPB	19 Oct 2022 14:17:34	U	SMPL
CRA - 1	CRA	973	0.2106	-	7470A	PPB	19 Oct 2022 14:19:51	U	SMPL
HighStd - 1	HighStd	45221	10.1417	-	7470A	PPB	19 Oct 2022 14:27:00	U	SMPL
ChkStd - 1	ChkStd	29957	6.7158	-	7470A	PPB	19 Oct 2022 14:29:15	U	SMPL
PB148436BL - 1	PBW	-17	-0.0116	-	7470A	PPB	19 Oct 2022 14:31:33	U	SMPL
PB148436BS - 1	LCSW	15303	3.4268	-	7470A	PPB	19 Oct 2022 14:33:50	U	SMPL
N4955-07 - 1	LOD-MDL-WATER-01-QT4-2022	279	0.0548	-	7470A	PPB	19 Oct 2022 14:36:07	U	SMPL
N4955-08 - 1	LOQ-WATER-02-QT4-2022	936	0.2023	-	7470A	PPB	19 Oct 2022 14:38:23	U	SMPL
N5176-01 - 1	5236-1	102	0.0151	-	7470A	PPB	19 Oct 2022 14:40:40	U	SMPL
N5176-01DUP - 1	5236-1DUP	130	0.0214	-	7470A	PPB	19 Oct 2022 14:42:57	U	SMPL
N5176-01MS - 1	5236-1MS	15631	3.5005	-	7470A	PPB	19 Oct 2022 14:45:13	U	SMPL
CCV49 - 1	CCV49	21037	4.7138	-	7470A	PPB	19 Oct 2022 14:47:28	U	SMPL
CCB49 - 1	CCB49	67	0.0072	-	7470A	PPB	19 Oct 2022 14:49:44	U	SMPL
N5176-01MSD - 1	5236-1MSD	15226	3.4096	-	7470A	PPB	19 Oct 2022 14:52:00	U	SMPL
N5176-02 - 1	5236-2	22	-0.0029	-	7470A	PPB	19 Oct 2022 14:54:15	U	SMPL
N5181-01 - 1	PIT-14-EXCAVATION-WATER	13217	2.9587	-	7470A	PPB	19 Oct 2022 14:56:32	U	SMPL
N5176-01LX5 - 1		-171	-0.0462	-	7470A	PPB	19 Oct 2022 14:58:48	U	SMPL
N5176-01A - 1		17538	3.9285	-	7470A	PPB	19 Oct 2022 15:01:05	U	SMPL
CCV50 - 1	CCV50	21706	4.8639	-	7470A	PPB	19 Oct 2022 15:03:21	U	SMPL
CCB50 - 1	CCB50	51	0.0037	-	7470A	PPB	19 Oct 2022 15:05:38	U	SMPL