



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122534

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/19/2022	14:30
2	CAL2	1	Water	NA	NA	20.7	7.00	10/19/2022	14:31
3	CAL3	1	Water	NA	NA	20.7	10.03	10/19/2022	14:32
4	ICV	1	Water	NA	NA	20.8	6.99	10/19/2022	14:35
5	CCV1	1	Water	NA	NA	20.7	2.01	10/19/2022	14:36
6	N5176-01	1	Water	NA	NA	20.3	7.53	10/19/2022	14:39
7	N5176-02	1	Water	NA	NA	20.4	7.81	10/19/2022	14:43
8	N5181-01	1	Water	NA	NA	20.4	6.92	10/19/2022	14:45
9	N5187-02	1	Water	NA	NA	20.1	6.24	10/19/2022	14:49
10	N5189-01	1	Water	NA	NA	20.6	7.66	10/19/2022	14:53
11	N5189-02	1	Water	NA	NA	20.7	8.04	10/19/2022	15:00
12	N5191-01	1	Water	NA	NA	20.5	5.65	10/19/2022	15:03
13	N5191-01DUP	1	Water	NA	NA	20.6	5.66	10/19/2022	15:04
14	CCV2	1	Water	NA	NA	20.8	12.03	10/19/2022	15:05