



PERCENT SOLID

Supervisor: Iwona  
Analyst: JIGNESH  
Date: 10/21/2022

OVENTEMP IN Celsius(°C): 108  
Time IN: 14:00  
In Date: 10/20/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:10  
Out Date: 10/21/2022  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB122544

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|-----------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| N5205-01 | MH0AE5          | 1      | 0.91           | 9.05         | 9.96                    | 9.95                      | 99.9    |          |
| N5205-02 | MH0AE7          | 2      | 0.89           | 9.03         | 9.92                    | 8.9                       | 88.7    |          |
| N5205-03 | MH0AE7D         | 3      | 0.89           | 9.03         | 9.92                    | 8.9                       | 88.7    |          |
| N5205-04 | MH0AE7S         | 4      | 0.89           | 9.03         | 9.92                    | 8.9                       | 88.7    |          |
| N5205-05 | MH0AE8          | 5      | 0.89           | 9.01         | 9.9                     | 8.88                      | 88.7    |          |
| N5205-06 | MH0AE9          | 6      | 0.9            | 9.06         | 9.96                    | 9.67                      | 96.8    |          |
| N5205-07 | MH0AF0          | 7      | 0.88           | 9.08         | 9.96                    | 9.67                      | 96.8    |          |
| N5205-08 | MH0AF1          | 8      | 0.88           | 8.99         | 9.87                    | 9.63                      | 97.3    |          |
| N5205-09 | MH0AF2          | 9      | 0.88           | 8.99         | 9.87                    | 9.3                       | 93.7    |          |
| N5205-10 | MH0AF3          | 10     | 0.89           | 9.04         | 9.93                    | 9.66                      | 97.0    |          |
| N5205-11 | MH0AF5          | 11     | 0.91           | 8.92         | 9.83                    | 9.37                      | 94.8    |          |
| N5205-12 | MH0AF6          | 12     | 0.89           | 9.08         | 9.97                    | 9.74                      | 97.5    |          |
| N5205-13 | MH0AF7          | 13     | 0.88           | 9.07         | 9.95                    | 9.8                       | 98.3    |          |
| N5205-14 | MH0AF8          | 14     | 0.87           | 8.95         | 9.82                    | 9.62                      | 97.8    |          |
| N5205-15 | MH0AF9          | 15     | 0.87           | 8.97         | 9.84                    | 9.5                       | 96.2    |          |
| N5205-16 | MH0AG0          | 16     | 0.91           | 8.95         | 9.86                    | 9.07                      | 91.2    |          |
| N5205-17 | MH0AG1          | 17     | 0.89           | 9.06         | 9.95                    | 9.37                      | 93.6    |          |
| N5205-18 | MH0AG2          | 18     | 0.88           | 9.04         | 9.92                    | 9.33                      | 93.5    |          |
| N5205-19 | MH0AG3          | 19     | 0.87           | 9.05         | 9.92                    | 9.00                      | 89.8    |          |
| N5205-20 | MH0AG4          | 20     | 0.87           | 9.07         | 9.94                    | 8.99                      | 89.5    |          |
| N5205-21 | MH0AG5          | 21     | 0.89           | 9.05         | 9.94                    | 9.26                      | 92.5    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$