



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Jignesh

Parameter: Flash Point

Supervisor Review By: Sohil

Run Number: LB122557

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	10/21/2022	10:19
2	N5187-02		1	14	100.00	>212.0	>212.0	10/21/2022	11:00
3	N5187-02DUP		1	13	100.00	>212.0	>212.0	10/21/2022	11:30
4	N5220-03		1	13	100.00	>212.0	>212.0	10/21/2022	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 122557

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Flash-1021

WorkList ID : 164266

Department : Wet-Chemistry

Date : 10-21-2022 10:15:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/26/2022	Water	N5187-02	Flash Point	Cool 4 deg C	PSEG03	D11	BUR-1226	10/19/2022	1010B
10/27/2022	Water	N5220-03	Flash Point	Cool 4 deg C	PSEG03	D11	D-178	10/20/2022	1010B

Date/Time 10/21/2022 10:15
 Raw Sample Received by: ST (WC)
 Raw Sample Relinquished by: CP(SM)

Date/Time 10/21/2022 12:30
 Raw Sample Received by: CP(SM)
 Raw Sample Relinquished by: ST(WC)