



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/24/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/21/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 10/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122558

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5217-01	SB-A1-3.5-4.0	1	0.92	8.98	9.9	8.84	88.2	
N5217-02	SB-A1-COMP	2	0.93	8.58	9.51	8.27	85.5	
N5217-03	SB-B1-3.5-4.0	3	0.93	8.90	9.83	8.72	87.5	
N5217-04	SB-B1-COMP	4	0.91	8.97	9.88	8.87	88.7	
N5217-05	SB-G1-3.5-4.0	5	0.91	8.71	9.62	9.18	94.9	
N5217-06	SB-G1-COMP	6	0.91	8.77	9.68	9.33	96.0	
N5217-07	SB-D1-3.5-4.0	7	0.92	8.79	9.71	8.27	83.6	
N5217-08	SB-D1-COMP	8	0.96	9.01	9.97	9.34	93.0	
N5217-09	SB-I1-3.5-4.0	9	0.93	8.76	9.69	8.93	91.3	
N5217-10	SB-I1-COMP	10	0.94	8.99	9.93	9.19	91.8	
N5217-11	SB-M1-3.5-4.0	11	0.95	8.60	9.55	8.76	90.8	
N5217-12	SB-M1-COMP	12	0.95	8.75	9.7	8.29	83.9	
N5217-13	SB-L1-3.5-4.0	13	0.96	8.96	9.92	8.03	78.9	
N5217-14	SB-L1-COMP	14	0.91	8.98	9.89	8.81	88.0	
N5217-15	SB-K1-3.5-4.0	15	0.93	8.88	9.81	8.8	88.6	
N5217-16	SB-K1-COMP	16	0.92	8.98	9.9	8.89	88.8	
N5217-17	SB-C1-7.5-8.0	17	0.95	8.97	9.92	8.44	83.5	
N5217-18	SB-C1-COMP	18	0.93	9.03	9.96	9.64	96.5	
N5217-19	SB-E1-7.5-8.0	19	0.91	8.90	9.81	9.09	91.9	
N5217-20	SB-E1-COMP	20	0.98	8.89	9.87	8.44	83.9	
N5217-21	SB-E2-7.5-8.0	21	0.95	9.01	9.96	7.95	77.7	
N5217-22	SB-E2-COMP	22	0.93	8.84	9.77	7.59	75.3	
N5217-23	SB-E3-7.5-8.0	23	0.93	9.05	9.98	9.39	93.5	
N5217-24	SB-E3-COMP	24	0.91	8.69	9.6	7.84	79.7	
N5218-01	SP-1	25	0.91	8.68	9.59	8.94	92.5	
N5218-03	SP-1A	26	0.92	8.81	9.73	9.1	92.8	
N5218-04	SP-1B	27	0.93	8.65	9.58	8.92	92.4	
N5218-05	SP-1C	28	0.91	9.07	9.98	9.42	93.8	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/24/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/21/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 10/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122558

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5218-06	SP-1D	29	0.93	8.55	9.48	8.72	91.1	
N5218-07	SP-1E	30	0.93	8.68	9.61	9.2	95.3	
N5218-08	SP-1F	31	0.93	8.59	9.52	9.06	94.6	
N5218-09	SP-1G	32	0.93	8.68	9.61	8.95	92.4	
N5218-10	SP-1H	33	0.93	8.93	9.86	9.14	91.9	
N5218-11	SP-2	34	0.93	8.69	9.62	8.89	91.6	
N5218-13	SP-2A	35	0.91	8.88	9.79	9.02	91.3	
N5218-14	SP-2B	36	0.91	8.83	9.74	8.8	89.4	
N5218-15	SP-2C	37	0.91	9.06	9.97	9.13	90.7	
N5218-16	SP-2D	38	0.93	8.82	9.75	8.99	91.4	
N5218-17	SP-2E	39	0.93	9.05	9.98	9.51	94.8	
N5218-18	SP-2F	40	0.92	8.90	9.82	9.04	91.2	
N5218-19	SP-2G	41	0.91	8.99	9.9	9.15	91.7	
N5218-20	SP-2H	42	0.92	8.96	9.88	8.74	87.3	
N5218-21	SP-3	43	0.93	8.59	9.52	9.00	93.9	
N5218-23	SP-3A	44	0.92	8.74	9.66	8.84	90.6	
N5218-24	SP-3B	45	0.93	8.84	9.77	8.99	91.2	
N5218-25	SP-3C	46	0.95	8.71	9.66	9.08	93.3	
N5218-26	SP-3D	47	0.9	8.84	9.74	9.00	91.6	
N5218-27	SP-3E	48	0.93	8.80	9.73	8.94	91.0	
N5218-28	SP-3F	49	0.9	9.08	9.98	9.15	90.9	
N5218-29	SP-3G	50	0.91	8.79	9.7	8.69	88.5	
N5218-30	SP-3H	51	0.91	8.88	9.79	9.08	92.0	
N5218-31	SP-4	52	0.93	8.71	9.64	9.05	93.2	
N5218-33	SP-4A	53	0.9	8.75	9.65	8.96	92.1	
N5218-34	SP-4B	54	0.93	8.82	9.75	8.94	90.8	
N5218-35	SP-4C	55	0.89	8.98	9.87	9.24	93.0	
N5218-36	SP-4D	56	0.92	9.06	9.98	9.39	93.5	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/24/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/21/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 10/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122558

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5218-37	SP-4E	57	0.9	8.82	9.72	9.16	93.7	
N5218-38	SP-4F	58	0.9	8.93	9.83	9.18	92.7	
N5218-39	SP-4G	59	0.92	8.70	9.62	8.99	92.8	
N5218-40	SP-4H	60	0.89	8.88	9.77	9.13	92.8	
N5221-01	CS-1- (8-POINT-COMPOSITE) -PI LE-1	61	0.9	9.08	9.98	9.55	95.3	
N5221-02	CS-2- (5-POINT-COMPOSITE) -PI LE-2	62	0.89	8.82	9.71	8.74	89.0	
N5221-03	CS-3- (8-POINT-COMPOSITE) -PI LE-3	63	0.89	9.06	9.95	9.07	90.3	
N5221-04	CS-4- (8-POINT-COMPOSITE) -PI LE-4	64	0.88	8.78	9.66	9.2	94.8	
N5224-01	SOILPILE-MIX	65	0.9	8.84	9.74	9.41	96.3	
N5224-02	SOILPILE-MIX-EPH-2	66	0.9	8.84	9.74	9.41	96.3	
N5229-01	LRH-COMP-1	67	0.89	8.85	9.74	8.69	88.1	
N5229-02	LRH-EPH-1	68	0.9	8.70	9.6	8.78	90.6	
N5229-03	LRH-COMP-2	69	0.92	8.96	9.88	9.46	95.3	
N5229-04	LRH-EPH-2	70	0.92	9.05	9.97	8.89	88.1	
N5234-01	LSP-SS-03-01-03	84	0.92	9.04	9.96	9.34	93.1	
N5234-02	LSP-SS-03-03-05	85	0.92	8.98	9.9	8.67	86.3	
N5234-03	LSP-SS-03-01-03-D	86	0.93	8.79	9.72	8.87	90.3	
N5234-04	LSP-SS-04-05-15	87	0.93	8.85	9.78	8.91	90.2	
N5234-05	LSP-SS-04-15-3	88	0.93	8.88	9.81	7.99	79.5	
N5234-06	LSP-SS-06-00-20	89	0.91	8.79	9.7	9.12	93.4	
N5234-07	LSP-SS-06-20-40	90	0.95	9.00	9.95	9.34	93.2	
N5234-08	LSP-SS-05-08-13	91	0.91	9.03	9.94	8.39	82.8	
N5234-09	LSP-SS-07-00-20	92	0.93	9.04	9.97	8.99	89.2	
N5234-10	LSP-SS-07-20-40	93	0.91	8.66	9.57	8.84	91.6	
N5235-01	LSP-SS-01-00-20	94	0.93	8.91	9.84	9.16	92.4	



PERCENT SOLID

Supervisor: Iwona
Analyst: JIGNESH
Date: 10/24/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/21/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:30
Out Date: 10/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB122558

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N5235-02	N5235-01MS	95	0.93	8.91	9.84	9.16	92.4	
N5235-03	N5235-01MSD	96	0.93	8.91	9.84	9.16	92.4	
N5235-04	LSP-DS-01-09-11	97	0.93	9.05	9.98	9.55	95.2	
N5235-05	LSP-SS-02-01-03	98	0.96	9.01	9.97	9.46	94.3	
N5235-06	LSP-SS-02-03-05	99	0.92	9.05	9.97	9.55	95.4	
N5238-01	CHRT-20430	100	0.92	8.88	9.8	8.85	89.3	
N5238-02	CHRT-20430-EPH-2	101	0.92	9.04	9.96	9.22	91.8	
N5238-03	2673	102	0.92	9.05	9.97	9.01	89.4	
N5238-04	2673-EPH-2	103	0.91	9.06	9.97	9.3	92.6	
N5238-05	RBR-251375	104	0.91	8.97	9.88	9.09	91.2	
N5238-06	RBR-251375-EPH-2	105	0.96	9.00	9.96	9.26	92.2	
N5239-01	34252	71	0.92	8.65	9.57	9.00	93.4	
N5239-02	34250	72	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5239-03	34251	73	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N5253-01	164718	106	0.92	8.62	9.54	7.2	72.9	GEL MATRIX
N5253-03	124055	107	0.94	9.00	9.94	5.8	54.0	GEL MATRIX
N5253-05	124042	108	0.89	8.77	9.66	7.13	71.2	GEL MATRIX
N5253-07	124019	109	0.96	9.01	9.97	7.18	69.0	GEL MATRIX
N5254-01	WC-1-102122	74	0.91	8.74	9.65	8.16	83.0	
N5254-02	WC-1-102122-EPH-2	75	0.91	8.74	9.65	8.16	83.0	
N5254-04	WC-2-102122	76	0.96	8.99	9.95	8.58	84.8	
N5254-05	WC-2-102122-EPH-2	77	0.96	8.99	9.95	8.58	84.8	
N5254-07	WC-3-102122	78	0.91	8.86	9.77	8.00	80.0	
N5254-08	WC-3-102122-EPH-2	79	0.91	8.86	9.77	8.00	80.0	
N5254-10	WC-4-102122	80	0.93	9.02	9.95	8.14	79.9	
N5254-11	WC-4-102122-EPH-2	81	0.93	9.02	9.95	8.14	79.9	
N5254-13	WC-5-102122	82	0.91	8.96	9.87	8.3	82.5	
N5254-14	WC-5-102122-EPH-2	83	0.91	8.96	9.87	8.3	82.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$