



TOTAL VOLATILE SOLIDS 160.4

TEMP1 IN: 103 °C 10/20/2022 16:25 TEMP1 OUT: 103 °C 10/21/2022 08:00
TEMP2 IN: 103 °C 10/21/2022 08:30 TEMP2 OUT: 104 °C 10/21/2022 10:00
TEMP3 IN: 540 °C 10/21/2022 10:30 TEMP3 OUT: 540 °C 10/21/2022 12:00
TEMP4 IN: 560 °C 10/21/2022 12:30 TEMP4 OUT: 560 °C 10/21/2022 14:00

Run Number: LB122561

SUPERVISOR: Iwona

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB122561BL	89.6784	89.6784	100	89.6784	89.6784	89.6784	89.6784	0.0000	0
2	N4788-33	101.0177	101.0177	100	101.0360	101.036	101.0180	101.0180	0.0180	180

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

VB122561

WorkList Name : TVS N47788-1

WorkList ID : 164273

Department : Wet-Chemistry

Date : 10-20-2022 12:14:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
09/26/2022	Water	N4788-33	TVS	Cool 4 deg C	CHEM02	QA Of	PT-VSOL-WP	09/19/2022	160.4

Date/Time 10/20/22 12:130
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)

Date/Time 10/20/22 16:30
 Raw Sample Received by: JP (WC)
 Raw Sample Relinquished by: JP (WC)