

LB122T65


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

50172



A= 0.0000e+000

B= 2.0096e-004

C= 2.8500e-002

Rho= 0.9997837

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.038	0.038	49	0.000	49					
0.2	0.200	0.211	0.011	910	0.0 %	910					
2.5	2.500	2.491	-0.009	12256	0.0 %	12256					
5.0	5.000	4.994	-0.006	24709	0.0 %	24709					
7.5	7.500	7.354	-0.146	36453	0.0 %	36453					
10.0	10.000	10.111	0.111	50172	0.0 %	50172					

0/0)
150051-200051

LB122565 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Oct 2022 13:06:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	49	-	0.0000	7470A	PPB	21 Oct 2022 13:07:46	S	Std
0.2 - 1	SD	910	-	0.2000	7470A	PPB	21 Oct 2022 13:10:04	S	Std
2.5 - 1	SD	12256	-	2.5000	7470A	PPB	21 Oct 2022 13:12:21	S	Std
5.0 - 1	SD	24709	-	5.0000	7470A	PPB	21 Oct 2022 13:17:25	S	Std
7.5 - 1	SD	36453	-	7.5000	7470A	PPB	21 Oct 2022 13:24:46	S	Std
10.0 - 1	SD	50172	-	10.0000	7470A	PPB	21 Oct 2022 13:29:34	S	Std
ICV86 - 1	ICV86	21329	4.3147	-	7470A	PPB	21 Oct 2022 13:37:37	U	SMPL
ICB86 - 1	ICB86	64	0.0414	-	7470A	PPB	21 Oct 2022 13:42:08	U	SMPL
CCV58 - 1	CCV58	24788	5.0098	-	7470A	PPB	21 Oct 2022 13:51:28	U	SMPL
CCB58 - 1	CCB58	-22	0.0241	-	7470A	PPB	21 Oct 2022 13:53:43	U	SMPL
CRA - 1	CRA	968	0.2230	-	7470A	PPB	21 Oct 2022 13:55:58	U	SMPL
HighStd - 1	HighStd	50655	10.2080	-	7470A	PPB	21 Oct 2022 13:58:13	U	SMPL
ChkStd - 1	ChkStd	33848	6.8305	-	7470A	PPB	21 Oct 2022 14:00:29	U	SMPL
PB148405BL - 1	PBW	-52	0.0181	-	7470A	PPB	21 Oct 2022 14:02:46	U	SMPL
PB148405BS - 1	LCSW	20456	4.1393	-	7470A	PPB	21 Oct 2022 14:05:02	U	SMPL
N5158-01 - 1	COMP-TOTE	47	0.0379	-	7470A	PPB	21 Oct 2022 14:07:18	U	SMPL
N5158-01DUP - 1	COMP-TOTEDUP	93	0.0472	-	7470A	PPB	21 Oct 2022 14:09:35	U	SMPL
N5158-01MS - 1	COMP-TOTEMS	19099	3.8666	-	7470A	PPB	21 Oct 2022 14:11:52	U	SMPL
N5158-01MSD - 1	COMP-TOTEMSD	18754	3.7972	-	7470A	PPB	21 Oct 2022 14:14:07	U	SMPL
PB148481BL - 1	PBW	28	0.0341	-	7470A	PPB	21 Oct 2022 14:16:23	U	SMPL
CCV59 - 1	CCV59	24735	4.9992	-	7470A	PPB	21 Oct 2022 14:18:38	U	SMPL
CCB59 - 1	CCB59	-22	0.0241	-	7470A	PPB	21 Oct 2022 14:20:54	U	SMPL
PB148481BS - 1	LCSW	21521	4.3533	-	7470A	PPB	21 Oct 2022 14:23:09	U	SMPL
N5183-02 - 1	PIT-14-EXCAVATION-SOIL	9	0.0303	-	7470A	PPB	21 Oct 2022 14:25:24	U	SMPL
N5186-02 - 1	TEXAS-SP-1019	63	0.0412	-	7470A	PPB	21 Oct 2022 14:27:40	U	SMPL
N5204-01 - 1	20071	61	0.0408	-	7470A	PPB	21 Oct 2022 14:29:56	U	SMPL
N5204-01DUP - 1	20071DUP	61	0.0408	-	7470A	PPB	21 Oct 2022 14:32:13	U	SMPL
N5204-01MS - 1	20071MS	19725	3.9924	-	7470A	PPB	21 Oct 2022 14:34:29	U	SMPL
N5204-01MSD - 1	20071MSD	19258	3.8985	-	7470A	PPB	21 Oct 2022 14:36:46	U	SMPL
N5207-01 - 1	2689	-12	0.0261	-	7470A	PPB	21 Oct 2022 14:39:03	U	SMPL
CCV60 - 1	CCV60	24173	4.8862	-	7470A	PPB	21 Oct 2022 14:41:18	U	SMPL
CCB60 - 1	CCB60	30	0.0345	-	7470A	PPB	21 Oct 2022 14:43:34	U	SMPL
PB148418TB - 1	PB148418TB	84	0.0454	-	7470A	PPB	21 Oct 2022 14:45:49	U	SMPL
PB148420TB - 1	PB148420TB	80	0.0446	-	7470A	PPB	21 Oct 2022 14:48:05	U	SMPL
N5158-01LX5 - 1		-3	0.0279	-	7470A	PPB	21 Oct 2022 14:50:21	U	SMPL
N5158-01A - 1		18947	3.8360	-	7470A	PPB	21 Oct 2022 14:52:36	U	SMPL
N5204-01LX5 - 1		-79	0.0126	-	7470A	PPB	21 Oct 2022 14:54:52	U	SMPL
N5204-01A - 1		20930	4.2345	-	7470A	PPB	21 Oct 2022 14:57:07	U	SMPL
CCV61 - 1	CCV61	25595	5.1720	-	7470A	PPB	21 Oct 2022 14:59:24	U	SMPL
CCB61 - 1	CCB61	-39	0.0207	-	7470A	PPB	21 Oct 2022 15:01:40	U	SMPL