



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122569

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/21/2022	13:43
2	CAL2	1	Water	NA	NA	20.7	7.00	10/21/2022	13:44
3	CAL3	1	Water	NA	NA	20.7	10.03	10/21/2022	13:45
4	ICV	1	Water	NA	NA	20.6	6.98	10/21/2022	13:46
5	CCV1	1	Water	NA	NA	20.7	2.01	10/21/2022	13:50
6	N5220-03	1	Water	NA	NA	20.9	7.51	10/21/2022	14:00
7	N5220-03DUP	1	Water	NA	NA	20.8	7.52	10/21/2022	14:01
8	N5239-04	1	Water	NA	NA	20.1	7.47	10/21/2022	14:05
9	CCV2	1	Water	NA	NA	20.7	12.03	10/21/2022	14:10

WORKLIST(Hardcopy Internal Chain)

VB 122569

WorkList Name : ph n5239 WorkList ID : 164275 Department : Wet-Chemistry Date : 10-21-2022 13:31:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/20/2022	Water	N5220-03	pH	Cool 4 deg C	PSEG03	D11	D-178	10/20/2022	9040C
10/21/2022	Water	N5239-04	pH	Cool 4 deg C	PSEG03	E11	34159	10/21/2022	9040C

Date/Time 10/21/22 13:35
 Raw Sample Received by: J.P. (W.C.)
 Raw Sample Relinquished by: Cif Sm

Date/Time 10/21/22 15:00
 Raw Sample Received by: Cif Sm
 Raw Sample Relinquished by: J.P. (W.C.)