



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB122570

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/21/2022	13:43
2	CAL2	1	Water	NA	NA	20.7	7.00	10/21/2022	13:44
3	CAL3	1	Water	NA	NA	20.7	10.03	10/21/2022	13:45
4	ICV	1	Water	NA	NA	20.6	6.98	10/21/2022	13:46
5	CCV1	1	Water	NA	NA	20.7	2.01	10/21/2022	13:50
6	N5239-01	1	Solid	20.01	20	20.7	5.75	10/21/2022	14:15
7	N5239-01DUP	1	Solid	20.02	20	20.6	5.76	10/21/2022	14:16
8	CCV2	1	Water	NA	NA	20.7	12.03	10/21/2022	14:19

U7122570

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s n5239 WorkList ID : 164276 Department : Wet-Chemistry Date : 10-21-2022 13:32:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/21/2022	Solid	N5239-01	pH	Cool 4 deg C	PSEG03	E11	34252	10/21/2022	9045D

Date/Time 10/21/22 13:37
Raw Sample Received by: 7D(WC)
Raw Sample Relinquished by: CP SM

Date/Time 10/21/22 15:10
Raw Sample Received by: CP SM
Raw Sample Relinquished by: 7D WC