


LB122574

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
B= 1.6174e-004
C= -6.7833e-002
Rho= 0.9998863
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.007	-0.007	375	0.000	375					
0.2	0.200	0.189	-0.011	1585	0.0 %	1585					-6
2.5	2.500	2.493	-0.007	15834	0.0 %	15834					0
5.0	5.000	4.996	-0.004	31309	0.0 %	31309					0
7.5	7.500	7.609	0.109	47465	0.0 %	47465					+1
10.0	10.000	9.920	-0.080	61752	0.0 %	61752					-1

LB122574 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Oct 2022 15:57:35

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	375	-	0.0000	7471B	PPB	21 Oct 2022 15:57:43	S	Std
0.2 - 1	50.2	1585	-	0.2000	7471B	PPB	21 Oct 2022 16:00:00	S	Std
2.5 - 1	52.5	15834	-	2.5000	7471B	PPB	21 Oct 2022 16:02:17	S	Std
5.0 - 1	55	31309	-	5.0000	7471B	PPB	21 Oct 2022 16:04:34	S	Std
7.5 - 1	57.5	47465	-	7.5000	7471B	PPB	21 Oct 2022 16:06:51	S	Std
10.0 - 1	510	61752	-	10.0000	7471B	PPB	21 Oct 2022 16:09:08	S	Std
ICV87 - 1	ICV87	25653	4.0813	-	7471B	PPB	21 Oct 2022 16:12:02	U	SMPL
ICB87 - 1	ICB87	161	-0.0418	-	7471B	PPB	21 Oct 2022 16:14:17	U	SMPL
CCV62 - 1	CCV62	31438	5.0170	-	7471B	PPB	21 Oct 2022 16:16:34	U	SMPL
CCB62 - 1	CCB62	149	-0.0437	-	7471B	PPB	21 Oct 2022 16:18:49	U	SMPL
CRA - 1	CRA	1600	0.1910	-	7471B	PPB	21 Oct 2022 16:21:06	U	SMPL
HighStd - 1	HighStd	62088	9.9744	-	7471B	PPB	21 Oct 2022 16:23:21	U	SMPL
ChkStd - 1	ChkStd	44848	7.1860	-	7471B	PPB	21 Oct 2022 16:25:36	U	SMPL
PB148461BL - 1	PBS	214	-0.0332	-	7471B	PPB	21 Oct 2022 16:27:54	U	SMPL
PB148461BS - 1	LCSS	26435	4.2078	-	7471B	PPB	21 Oct 2022 16:30:12	U	SMPL
N4955-01 - 1	LOD-MDL-SOIL-01-QT4-2022	648	0.0370	-	7471B	PPB	21 Oct 2022 16:37:34	U	SMPL
N4955-01RE - 1	LOD-MDL-SOIL-01-QT4-2022RE	1257	0.1355	-	7471B	PPB	21 Oct 2022 16:39:50	U	SMPL
N4955-03 - 1	MDL-SOIL-03-QT4-2022	779	0.0582	-	7471B	PPB	21 Oct 2022 16:42:07	U	SMPL
N5186-01 - 1	TEXAS-SP-1019	1185	0.1238	-	7471B	PPB	21 Oct 2022 16:44:24	U	SMPL
N5186-01DUP - 1	TEXAS-SP-1019DUP	949	0.0857	-	7471B	PPB	21 Oct 2022 16:46:40	U	SMPL
CCV63 - 1	CCV63	31244	4.9856	-	7471B	PPB	21 Oct 2022 16:48:55	U	SMPL
CCB63 - 1	CCB63	132	-0.0465	-	7471B	PPB	21 Oct 2022 16:51:11	U	SMPL
N5186-01MS - 1	TEXAS-SP-1019MS	25338	4.0304	-	7471B	PPB	21 Oct 2022 16:53:27	U	SMPL
N5186-01MSD - 1	TEXAS-SP-1019MSD	23591	3.7478	-	7471B	PPB	21 Oct 2022 16:55:43	U	SMPL
N5187-01 - 1	BUR-1202	616	0.0318	-	7471B	PPB	21 Oct 2022 16:58:00	U	SMPL
N5202-01 - 1	SU-04-101922	2381	0.3173	-	7471B	PPB	21 Oct 2022 17:00:17	U	SMPL
N5207-02 - 1	VNJ-221	4047	0.5867	-	7471B	PPB	21 Oct 2022 17:02:33	U	SMPL
N5209-01 - 1	CL-02-101922	2447	0.3279	-	7471B	PPB	21 Oct 2022 17:04:51	U	SMPL
N5186-01LX5 - 1		746	0.0528	-	7471B	PPB	21 Oct 2022 17:07:07	U	SMPL
N5186-01A - 1		26448	4.2099	-	7471B	PPB	21 Oct 2022 17:09:24	U	SMPL
CCV64 - 1	CCV64	31432	5.0160	-	7471B	PPB	21 Oct 2022 17:11:41	U	SMPL
CCB64 - 1	CCB64	121	-0.0483	-	7471B	PPB	21 Oct 2022 17:13:57	U	SMPL