



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB122576

BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP99524
calibration std. phosphate 0.5 ppm	WP99523
calibration std. phosphate 0.3 ppm	WP99522
calibration std. phosphate 0.1 ppm	WP99521
calibration std. phosphate 0.05 ppm	WP99520
calibration std. 0 ppm	WP99519
phosphate CCV std.	WP99525
Combined reagent	WP99528
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP99526
Phosphate LOD-MDL Std 0.025ppm	WP99530

Intercept: -0.0011

Slope: 0.6537

Regression: 0.999964

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		10/21/2022	11:30
2	CAL2	0.05	1	50	50	0.033	0.052	-4	10/21/2022	11:30
3	CAL3	0.10	1	50	50	0.065	0.101	-1	10/21/2022	11:31
4	CAL4	0.30	1	50	50	0.191	0.294	2	10/21/2022	11:31
5	CAL5	0.50	1	50	50	0.325	0.499	0.2	10/21/2022	11:32
6	CAL6	1.00	1	50	50	0.654	1.002	-0.2	10/21/2022	11:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: Jignesh

Run Number: LB122576

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.493	10/21/2022	11:33
2	ICB		1	50	50	0.002	0.005	10/21/2022	11:33
3	CCV1	0.50	1	50	50	0.322	0.494	10/21/2022	11:45
4	CCB1		1	50	50	0.002	0.005	10/21/2022	11:50
5	PB148499BL		1	1.00	50	0.003	0.006	10/21/2022	11:51
6	PB148499BS	0.50	1	1.00	50	0.321	0.493	10/21/2022	11:51
7	N4955-03		1	1.00	50	0.016	0.026	10/21/2022	11:52
8	CCV2	0.50	1	50	50	0.326	0.500	10/21/2022	11:52
9	CCB2		1	50	50	0.002	0.005	10/21/2022	11:53