

LB122582


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

52202



A= 0.0000e+000

B= 1.9218e-004

C= 9.8933e-003

Rho= 0.9998875

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.043	0.043	172	0.000	172					
0.2	0.200	0.218	0.018	1084	0.0 %	1084					
2.5	2.500	2.493	-0.007	12919	0.0 %	12919					
5.0	5.000	4.882	-0.118	25352	0.0 %	25352					
7.5	7.500	7.522	0.022	39086	0.0 %	39086					
10.0	10.000	10.042	0.042	52202	0.0 %	52202					

0/01
1
9
0
2
0
0

LB122582 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Oct 2022 10:30:12

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	172	-	0.0000	7470A	PPB	24 Oct 2022 10:58:56	S	Std
0.2 - 1	50.2	1084	-	0.2000	7470A	PPB	24 Oct 2022 11:01:14	S	Std
2.5 - 1	52.5	12919	-	2.5000	7470A	PPB	24 Oct 2022 11:03:31	S	Std
5.0 - 1	55	25352	-	5.0000	7470A	PPB	24 Oct 2022 11:05:47	S	Std
7.5 - 1	57.5	39086	-	7.5000	7470A	PPB	24 Oct 2022 11:08:04	S	Std
10.0 - 1	511	52202	-	10.0000	7470A	PPB	24 Oct 2022 11:10:19	S	Std
ICV88 - 1	ICV88	20392	3.9289	-	7470A	PPB	24 Oct 2022 11:13:04	U	SMPL
ICB88 - 1	ICB88	-314	-0.0505	-	7470A	PPB	24 Oct 2022 11:15:20	U	SMPL
CCV65 - 1	CCV65	27295	5.2558	-	7470A	PPB	24 Oct 2022 11:17:35	U	SMPL
CCB65 - 1	CCB65	61	0.0218	-	7470A	PPB	24 Oct 2022 11:26:44	U	SMPL
CRA - 1	CRA	1105	0.2223	-	7470A	PPB	24 Oct 2022 11:28:59	U	SMPL
HighStd - 1	HighStd	53125	10.2197	-	7470A	PPB	24 Oct 2022 11:31:14	U	SMPL
ChkStd - 1	ChkStd	35068	6.7494	-	7470A	PPB	24 Oct 2022 11:33:29	U	SMPL
PB148459BL - 1	PBW	-134	-0.0159	-	7470A	PPB	24 Oct 2022 11:35:47	U	SMPL
PB148459BS - 1	LCSW	20253	3.9022	-	7470A	PPB	24 Oct 2022 11:38:03	U	SMPL
N4955-07 - 1	LOD-MDL-WATER-01-QT4-2022	179	0.0443	-	7470A	PPB	24 Oct 2022 11:40:20	U	SMPL
N4955-07RE - 1	LOD-MDL-WATER-01-QT4-2022RE	698	0.1440	-	7470A	PPB	24 Oct 2022 11:42:36	U	SMPL
N4955-09 - 1	MDL-WATER-03-QT4-2022	246	0.0572	-	7470A	PPB	24 Oct 2022 11:47:32	U	SMPL
N5185-03 - 1	MW-02	74	0.0241	-	7470A	PPB	24 Oct 2022 11:49:49	U	SMPL
N5185-05 - 1	MW-04	184	0.0453	-	7470A	PPB	24 Oct 2022 11:52:04	U	SMPL
CCV66 - 1	CCV66	26278	5.0801	-	7470A	PPB	24 Oct 2022 11:54:20	U	SMPL
CCB66 - 1	CCB66	-126	-0.0143	-	7470A	PPB	24 Oct 2022 11:56:35	U	SMPL
N5185-07 - 1	MW-05	154	0.0395	-	7470A	PPB	24 Oct 2022 11:58:51	U	SMPL
N5185-09 - 1	MW-08	338	0.0749	-	7470A	PPB	24 Oct 2022 12:01:06	U	SMPL
N5185-11 - 1	MW-10	310	0.0695	-	7470A	PPB	24 Oct 2022 12:03:21	U	SMPL
N5185-11DUP - 1	MW-10DUP	348	0.0768	-	7470A	PPB	24 Oct 2022 12:05:37	U	SMPL
N5185-13 - 1	N5185-11MS	19584	3.7736	-	7470A	PPB	24 Oct 2022 12:07:53	U	SMPL
N5185-14 - 1	N5185-11MSD	18346	3.5357	-	7470A	PPB	24 Oct 2022 12:10:09	U	SMPL
N5185-17 - 1	PZ-02	32	0.0160	-	7470A	PPB	24 Oct 2022 12:12:25	U	SMPL
N5185-19 - 1	PZ-3R	324	0.0722	-	7470A	PPB	24 Oct 2022 12:14:42	U	SMPL
N5185-21 - 1	RW-01	110	0.0310	-	7470A	PPB	24 Oct 2022 12:17:00	U	SMPL
N5185-23 - 1	RW-02	163	0.0412	-	7470A	PPB	24 Oct 2022 12:19:15	U	SMPL
CCV67 - 1	CCV67	26654	5.1324	-	7470A	PPB	24 Oct 2022 12:21:31	U	SMPL
CCB67 - 1	CCB67	-99	-0.0091	-	7470A	PPB	24 Oct 2022 12:23:46	U	SMPL
N5187-02 - 1	BUR-1226	448	0.0960	-	7470A	PPB	24 Oct 2022 12:26:02	U	SMPL
N5189-01 - 1	1431-1	275	0.0827	-	7470A	PPB	24 Oct 2022 12:28:17	U	SMPL
N5189-02 - 1	1431-2	242	0.0564	-	7470A	PPB	24 Oct 2022 12:30:33	U	SMPL
PB148530BL - 1	PBW	-54	-0.0005	-	7470A	PPB	24 Oct 2022 12:38:01	U	SMPL
PB148530BS - 1	LCSW	19892	3.8328	-	7470A	PPB	24 Oct 2022 12:40:17	U	SMPL
N5185-02 - 1	MW-01	502	0.1064	-	7470A	PPB	24 Oct 2022 12:42:33	U	SMPL
N5185-04 - 1	MW-02	550	0.1156	-	7470A	PPB	24 Oct 2022 12:47:41	U	SMPL
N5185-06 - 1	MW-04	592	0.1237	-	7470A	PPB	24 Oct 2022 12:49:58	U	SMPL
N5185-08 - 1	MW-05	573	0.1200	-	7470A	PPB	24 Oct 2022 12:52:15	U	SMPL
N5185-10 - 1	MW-08	599	0.1250	-	7470A	PPB	24 Oct 2022 12:54:30	U	SMPL
CCV68 - 1	CCV68	27815	5.3555	-	7470A	PPB	24 Oct 2022 12:56:46	U	SMPL
CCB68 - 1	CCB68	-96	-0.0086	-	7470A	PPB	24 Oct 2022 12:59:02	U	SMPL
N5185-12 - 1	MW-10	645	0.1339	-	7470A	PPB	24 Oct 2022 13:01:17	U	SMPL
N5185-12DUP - 1	MW-10DUP	611	0.1273	-	7470A	PPB	24 Oct 2022 13:03:33	U	SMPL
N5185-15 - 1	N5185-12MS	19676	3.7913	-	7470A	PPB	24 Oct 2022 13:05:48	U	SMPL
N5185-16 - 1	N5185-12MSD	18941	3.6501	-	7470A	PPB	24 Oct 2022 13:08:04	U	SMPL
N5185-18 - 1	PZ-02	489	0.1039	-	7470A	PPB	24 Oct 2022 13:10:21	U	SMPL
N5185-20 - 1	PZ-3R	910	0.1848	-	7470A	PPB	24 Oct 2022 13:12:37	U	SMPL
N5185-22 - 1	RW-01	672	0.1390	-	7470A	PPB	24 Oct 2022 13:14:53	U	SMPL
N5185-24 - 1	RW-02	603	0.1258	-	7470A	PPB	24 Oct 2022 13:17:09	U	SMPL
N5185-01 - 1	MW-01	120	0.0330	-	7470A	PPB	24 Oct 2022 13:19:25	U	SMPL
CCV69 - 1	CCV69	26481	5.0991	-	7470A	PPB	24 Oct 2022 13:21:41	U	SMPL
CCB69 - 1	CCB69	-85	-0.0064	-	7470A	PPB	24 Oct 2022 13:23:57	U	SMPL
N5185-11LX5 - 1		-1	0.0097	-	7470A	PPB	24 Oct 2022 13:26:13	U	SMPL
N5185-11A - 1		20304	3.9120	-	7470A	PPB	24 Oct 2022 13:28:29	U	SMPL
N5185-12LX5 - 1		8	0.0114	-	7470A	PPB	24 Oct 2022 13:30:45	U	SMPL
N5185-12A - 1		20739	3.9956	-	7470A	PPB	24 Oct 2022 13:32:61	U	SMPL
CCV70 - 1	CCV70	27147	5.2271	-	7470A	PPB	24 Oct 2022 13:34:77	U	SMPL
CCB70 - 1	CCB70	-87	-0.0068	-	7470A	PPB	24 Oct 2022 13:36:93	U	SMPL