

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: jigneshRun Number: LB122588Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP98799

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB122588BL		1	20.7	0.76	0.83	10/20/2022	11:30
2	LB122588BS	1413	1	20.6	1314.00	1434.56	10/20/2022	11:35
3	N5191-01		1	20.3	1.92	2.11	10/20/2022	11:40
4	N5191-01DUP		1	20.4	1.93	2.12	10/24/2022	11:45

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

LB122588

WorkList Name : CONDUCTANCE-102022 WorkList ID : 164225 Department : Wet-Chemistry Date : 10-20-2022 10:09:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2022	Water	N5191-01	Conductance	Cool 4 deg C	STER10	N11	CELL-A-POOL-WATER(DE-IONIZED)	10/13/2022	SM2510 B

Date/Time 10/20/22 12:00
 Raw Sample Received by: 12/20/22
 Raw Sample Relinquished by: 12/20/22

Date/Time 10/20/22 12:20
 Raw Sample Received by: 12/20/22
 Raw Sample Relinquished by: 12/20/22