



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122591

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP99568
Turbidity Calibration std, 0NTU	WP99567
Turbidity Calibration std, 20NTU	WP99562
Turbidity Calibration std, 40NTU	WP99561
Turbidity Calibration std, 1NTU	WP99565
Turbidity Calibration std, 5NTU	WP99564
Turbidity LOQ std, 1.0NTU	WP99569
Turbidity Calibration - CCV std, 10 NTU	WP99563
10 NTU Standard 500 ml	W2971

Intercept: 0.0523

Slope: 0.9255

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.144	0.10		10/24/2022	10:20
2	CAL2	1	1	1.150	1.19	-18.6	10/24/2022	10:24
3	CAL3	5	1	4.420	4.72	5.6	10/24/2022	10:27
4	CAL4	10	1	9.440	10.14	-1.4	10/24/2022	10:30
5	CAL5	20	1	18.300	19.72	1.4	10/24/2022	10:34
6	CAL6	40	1	37.200	40.14	-0.3	10/24/2022	10:37



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: jignesh

Run Number: LB122591

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.350	10.046	10/24/2022	10:40
2	ICB		1	0.138	0.093	10/24/2022	10:43
3	CCV1	10	1	9.360	10.057	10/24/2022	10:47
4	CCB1		1	0.148	0.103	10/24/2022	10:50
5	LB122591BL		1	0.154	0.110	10/24/2022	10:53
6	N4955-07		1	0.523	0.509	10/24/2022	10:56
7	N4955-08		1	1.148	1.184	10/24/2022	11:00
8	CCV2	10	1	9.230	9.916	10/24/2022	11:03
9	CCB2		1	0.144	0.099	10/24/2022	11:06