



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122602

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/25/2022	09:43
2	N5239-04		1	12	100.00	>212.0	>212.0	10/25/2022	10:15
3	N5239-04DUP		1	13	100.00	>212.0	>212.0	10/25/2022	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB122602

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint-1025

WorkList ID : 164360

Department : Wet-Chemistry

Date : 10-25-2022 09:33:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/28/2022	Water	N5239-04	Flash Point	Cool 4 deg C	PSEG03	E11	34159	10/21/2022	1010B

Date/Time 10/25/22 9:40
Raw Sample Received by: SG (wc)
Raw Sample Relinquished by: SG (wc)

Date/Time 10/25/22 11:30
Raw Sample Received by: CP (SM)
Raw Sample Relinquished by: SG (wc)