



TOTAL SUSPENDED SOLIDS - SM2540D

TEMP1 IN: 103 °C 10/24/2022 14:00 TEMP1 OUT: 103 °C 10/24/2022 15:00
TEMP2 IN: 103 °C 10/24/2022 15:30 TEMP2 OUT: 104 °C 10/24/2022 16:30
TEMP3 IN: 104 °C 10/25/2022 09:00 TEMP3 OUT: 104 °C 10/25/2022 10:30
TEMP4 IN: 104 °C 10/25/2022 11:00 TEMP4 OUT: 103 °C 10/25/2022 12:30

SUPERVISOR: Iwona
ANALYST: JIGNESH
Date: 10/24/2022
Run Number: LB122603
BalanceID: WC SC-6
OvenID: WC OVEN-1
FilterID: 17416528
ThermometerID: WET OVEN#1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB122603BL	LB122603BL	1.3697	1.3697	100	1.3698	1.3698	1.3698	0.0001	1
2	LB122603BS	LB122603BS	1.4967	1.4967	100	1.5498	1.5498	1.5498	0.0531	531
3	N5226-01	20221093-198-SITE-WATER	1.4263	1.4263	1000	1.4573	1.4573	1.4573	0.0310	31
4	N5226-02	20221094-199-SITE-WATER	1.4379	1.4379	1000	1.4668	1.4668	1.4668	0.0289	28.9
5	N5226-03	20221096-201-SITE-WATER	1.4309	1.4309	1000	1.4623	1.4623	1.4623	0.0314	31.4
6	N5226-04	20221097-202-SITE-WATER	1.4386	1.4386	1000	1.4675	1.4675	1.4675	0.0289	28.9
7	N5226-05	20221098-203-SITE-WATER	1.4372	1.4372	1000	1.4678	1.4678	1.4678	0.0306	30.6
8	N5226-06	20221101-PC-SITE-WATER	1.4297	1.4297	1000	1.4539	1.4539	1.4539	0.0242	24.2
9	N5266-01	20221095-200-SITE-WATER	1.4225	1.4225	1000	1.4498	1.4498	1.4498	0.0273	27.3
10	N5266-02	20221099-204-SITE-WATER	1.4106	1.4106	1000	1.4385	1.4385	1.4385	0.0279	27.9
11	N5266-03	20221100-204.1-SITE-WATER	1.4033	1.4033	500	1.4369	1.4369	1.4369	0.0336	67.2
12	N5266-03DUP	20221100-204.1-SITE-WATERDU	1.4079	1.4079	500	1.4412	1.4412	1.4412	0.0333	66.6



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: JIGNESH

Date: 10/24/2022

Run Number: LB122603

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 10/24/2022 14:00 TEMP1 OUT: 103 °C 10/24/2022 15:00
TEMP2 IN: 103 °C 10/24/2022 15:30 TEMP2 OUT: 104 °C 10/24/2022 16:30
TEMP3 IN: 104 °C 10/25/2022 09:00 TEMP3 OUT: 104 °C 10/25/2022 10:30
TEMP4 IN: 104 °C 10/25/2022 11:00 TEMP4 OUT: 103 °C 10/25/2022 12:30

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L

A = Sample Volume (ml)
B = Final Empty Dish Weight (g)
C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)
D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

122603

WorkList Name: TSS N5266

WorkList ID: 164349

Department: Wet-Chemistry

Date: 10-25-2022 08:22:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/25/2022	Water	N5226-01	TSS	Cool 4 deg C	AQUA04	E11	20221093-198-SITE-WATER	10/18/2022	SM2540 D
10/25/2022	Water	N5226-02	TSS	Cool 4 deg C	AQUA04	E11	20221094-199-SITE-WATER	10/18/2022	SM2540 D
10/26/2022	Water	N5226-03	TSS	Cool 4 deg C	AQUA04	E11	20221096-201-SITE-WATER	10/19/2022	SM2540 D
10/26/2022	Water	N5226-04	TSS	Cool 4 deg C	AQUA04	E11	20221097-202-SITE-WATER	10/19/2022	SM2540 D
10/26/2022	Water	N5226-05	TSS	Cool 4 deg C	AQUA04	E11	20221098-203-SITE-WATER	10/19/2022	SM2540 D
10/26/2022	Water	N5226-06	TSS	Cool 4 deg C	AQUA04	E11	20221101-PC-SITE-WATER	10/19/2022	SM2540 D
10/28/2022	Water	N5266-01	TSS	Cool 4 deg C	AQUA04	F11	20221095-200-SITE-WATER	10/21/2022	SM2540 D
10/28/2022	Water	N5266-02	TSS	Cool 4 deg C	AQUA04	F11	20221099-204-SITE-WATER	10/21/2022	SM2540 D
10/28/2022	Water	N5266-03	TSS	Cool 4 deg C	AQUA04	F11	20221100-204.1-SITE-WATER	10/21/2022	SM2540 D

Date/Time 10/25/22 08:25
 Raw Sample Received by: JP (OC)
 Raw Sample Relinquished by: JP (OC)

Date/Time 10/25/22 11:00
 Raw Sample Received by: JP (OC)
 Raw Sample Relinquished by: JP (OC)