

LB122606

7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

51813

A= 0.0000e+000

B= 1.9327e-004

C= 2.4352e-002

Rho= 0.9999427

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.037	0.037	66	0.000	66					
0.2	0.200	0.208	0.008	948	0.0 %	948					
2.5	2.500	2.442	-0.058	12508	0.0 %	12508					
5.0	5.000	5.025	0.025	25874	0.0 %	25874					
7.5	7.500	7.450	-0.050	38424	0.0 %	38424					
10.0	10.000	10.038	0.038	51813	0.0 %	51813					

4
-4
-2
-1
0

LB122606 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Oct 2022 12:06:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	66	-	0.0000	7470A	PPB	25 Oct 2022 12:09:12	S	Std
0.2 - 1	50.2	948	-	0.2000	7470A	PPB	25 Oct 2022 12:11:30	S	Std
2.5 - 1	52.5	12508	-	2.5000	7470A	PPB	25 Oct 2022 12:13:47	S	Std
5.0 - 1	55	25874	-	5.0000	7470A	PPB	25 Oct 2022 12:16:04	S	Std
7.5 - 1	57.5	38424	-	7.5000	7470A	PPB	25 Oct 2022 12:18:20	S	Std
10.0 - 1	57.0	51813	-	10.0000	7470A	PPB	25 Oct 2022 12:20:36	S	Std
ICV90 - 1	ICV90	19954	3.8808	-	7470A	PPB	25 Oct 2022 12:23:28	U	SMPL
ICB90 - 1	ICB90	-30	0.0186	-	7470A	PPB	25 Oct 2022 12:25:43	U	SMPL
CCV75 - 1	CCV75	25495	4.9517	-	7470A	PPB	25 Oct 2022 12:27:58	U	SMPL
CCB75 - 1	CCB75	12	0.0267	-	7470A	PPB	25 Oct 2022 12:30:13	U	SMPL
CRA - 1	CRA	914	0.2010	-	7470A	PPB	25 Oct 2022 12:32:28	U	SMPL
HighStd - 1	HighStd	50537	9.7915	-	7470A	PPB	25 Oct 2022 12:34:43	U	SMPL
ChkStd - 1	ChkStd	34345	6.6621	-	7470A	PPB	25 Oct 2022 12:36:58	U	SMPL
PB148543BL - 1	PBW	-81	0.0087	-	7470A	PPB	25 Oct 2022 12:39:15	U	SMPL
PB148543BS - 1	LCSW	20316	3.9508	-	7470A	PPB	25 Oct 2022 12:41:32	U	SMPL
N5211-02 - 1	SB-SOILPILE-1020	-42	0.0162	-	7470A	PPB	25 Oct 2022 12:43:48	U	SMPL
N5217-02 - 1	SB-A1-COMP	22	0.0286	-	7470A	PPB	25 Oct 2022 12:46:04	U	SMPL
N5217-04 - 1	SB-B1-COMP	42	0.0325	-	7470A	PPB	25 Oct 2022 12:48:22	U	SMPL
N5217-06 - 1	SB-G1-COMP	25	0.0292	-	7470A	PPB	25 Oct 2022 12:50:39	U	SMPL
N5217-08 - 1	SB-D1-COMP	67	0.0373	-	7470A	PPB	25 Oct 2022 12:52:55	U	SMPL
CCV76 - 1	CCV76	26467	5.1396	-	7470A	PPB	25 Oct 2022 12:55:10	U	SMPL
CCB76 - 1	CCB76	16	0.0274	-	7470A	PPB	25 Oct 2022 12:57:25	U	SMPL
N5217-10 - 1	SB-I1-COMP	39	0.0319	-	7470A	PPB	25 Oct 2022 12:59:41	U	SMPL
N5217-12 - 1	SB-M1-COMP	159	0.0551	-	7470A	PPB	25 Oct 2022 13:03:45	U	SMPL
N5217-14 - 1	SB-L1-COMP	56	0.0352	-	7470A	PPB	25 Oct 2022 13:06:00	U	SMPL
N5217-16 - 1	SB-K1-COMP	37	0.0315	-	7470A	PPB	25 Oct 2022 13:08:16	U	SMPL
N5217-18 - 1	SB-C1-COMP	47	0.0334	-	7470A	PPB	25 Oct 2022 13:10:32	U	SMPL
N5217-20 - 1	SB-E1-COMP	33	0.0307	-	7470A	PPB	25 Oct 2022 13:12:48	U	SMPL
N5217-22 - 1	SB-E2-COMP	22	0.0286	-	7470A	PPB	25 Oct 2022 13:16:58	U	SMPL
N5217-24 - 1	SB-E3-COMP	20	0.0282	-	7470A	PPB	25 Oct 2022 13:19:14	U	SMPL
N5218-02 - 1	SP-1	38	0.0317	-	7470A	PPB	25 Oct 2022 13:21:32	U	SMPL
N5218-12 - 1	SP-2	53	0.0346	-	7470A	PPB	25 Oct 2022 13:23:47	U	SMPL
CCV77 - 1	CCV77	26336	5.1143	-	7470A	PPB	25 Oct 2022 13:26:03	U	SMPL
CCB77 - 1	CCB77	0	0.0244	-	7470A	PPB	25 Oct 2022 13:28:18	U	SMPL
N5218-22 - 1	SP-3	44	0.0329	-	7470A	PPB	25 Oct 2022 13:30:34	U	SMPL
N5218-32 - 1	SP-4	65	0.0369	-	7470A	PPB	25 Oct 2022 13:32:50	U	SMPL
N5218-32DUP - 1	SP-4DUP	67	0.0373	-	7470A	PPB	25 Oct 2022 13:35:05	U	SMPL
N5218-32MS - 1	SP-4MS	19668	3.8255	-	7470A	PPB	25 Oct 2022 13:37:21	U	SMPL
N5218-32MSD - 1	SP-4MSD	18917	3.6804	-	7470A	PPB	25 Oct 2022 13:39:36	U	SMPL
PB148544BL - 1	PBW	-49	0.0149	-	7470A	PPB	25 Oct 2022 13:41:53	U	SMPL
PB148544BS - 1	LCSW	18870	3.6713	-	7470A	PPB	25 Oct 2022 13:49:15	U	SMPL
N5220-03 - 1	D-178	-2	0.0240	-	7470A	PPB	25 Oct 2022 13:52:11	U	SMPL
N5220-03DUP - 1	D-178DUP	41	0.0323	-	7470A	PPB	25 Oct 2022 13:54:27	U	SMPL
N5220-03MS - 1	D-178MS	17140	3.3370	-	7470A	PPB	25 Oct 2022 13:56:43	U	SMPL
CCV78 - 1	CCV78	25896	5.0292	-	7470A	PPB	25 Oct 2022 13:58:58	U	SMPL
CCB78 - 1	CCB78	-77	0.0095	-	7470A	PPB	25 Oct 2022 14:01:13	U	SMPL
N5220-03MSD - 1	D-178MSD	17982	3.4997	-	7470A	PPB	25 Oct 2022 14:13:21	U	SMPL
N5218-32LX5 - 1		-25	0.0195	-	7470A	PPB	25 Oct 2022 14:16:31	U	SMPL
N5218-32A - 1		18863	3.6700	-	7470A	PPB	25 Oct 2022 14:18:47	U	SMPL
N5220-03LX5 - 1		-112	0.0027	-	7470A	PPB	25 Oct 2022 14:21:03	U	SMPL
N5220-03A - 1		20578	4.0014	-	7470A	PPB	25 Oct 2022 14:27:12	U	SMPL
CCV82 - 1	CCV82	26344	5.1158	-	7470A	PPB	25 Oct 2022 14:31:14	U	SMPL
CCB82 - 1	CCB82	10	0.0263	-	7470A	PPB	25 Oct 2022 14:33:30	U	SMPL