

LB122615


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

49220



A= 0.0000e+000

B= 2.0283e-004

C= -1.5368e-002

Rho= 0.9999575

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/1/1
0.0	0.000	-0.006	-0.006	44	0.000	44					1
0.2	0.200	0.170	-0.030	912	0.0 %	912					-15
2.5	2.500	2.504	0.004	12423	0.0 %	12423					0
5.0	5.000	5.070	0.070	25072	0.0 %	25072					0
7.5	7.500	7.495	-0.005	37026	0.0 %	37026					0
10.0	10.000	9.968	-0.032	49220	0.0 %	49220					0

lb122615 instrument id : cv

Method: 7470A

Operator: Admin

Date of Analysis: 26 Oct 2022 09:33:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	44	-	0.0000	7470A	PPB	26 Oct 2022 09:34:05	S	Std
0.2 - 1	50.2	912	-	0.2000	7470A	PPB	26 Oct 2022 09:36:21	S	Std
2.5 - 1	52.5	12423	-	2.5000	7470A	PPB	26 Oct 2022 09:38:38	S	Std
5.0 - 1	55	25072	-	5.0000	7470A	PPB	26 Oct 2022 09:40:55	S	Std
7.5 - 1	57.5	37026	-	7.5000	7470A	PPB	26 Oct 2022 09:43:12	S	Std
10.0 - 1	59	49220	-	10.0000	7470A	PPB	26 Oct 2022 09:48:03	S	Std
ICV92 - 1	ICV92	19816	4.0039	-	7470A	PPB	26 Oct 2022 09:50:49	U	SMPL
ICB92 - 1	ICB92	-85	-0.0326	-	7470A	PPB	26 Oct 2022 09:53:04	U	SMPL
CCV80 - 1	CCV80	24671	4.9886	-	7470A	PPB	26 Oct 2022 09:55:19	U	SMPL
CCB80 - 1	CCB80	-88	-0.0332	-	7470A	PPB	26 Oct 2022 09:57:35	U	SMPL
CRA - 1	CRA	878	0.1627	-	7470A	PPB	26 Oct 2022 09:59:51	U	SMPL
HighStd - 1	HighStd	49076	9.9387	-	7470A	PPB	26 Oct 2022 10:02:06	U	SMPL
ChkStd - 1	ChkStd	34090	6.8991	-	7470A	PPB	26 Oct 2022 10:04:22	U	SMPL
PB148570BL - 1	PBW	-90	-0.0336	-	7470A	PPB	26 Oct 2022 10:06:39	U	SMPL
PB148570BS - 1	LCSW	19154	3.8696	-	7470A	PPB	26 Oct 2022 10:08:57	U	SMPL
N5252-01 - 1	001-PPE-DRUM	-46	-0.0247	-	7470A	PPB	26 Oct 2022 10:11:13	U	SMPL
N5252-02 - 1	002-POLY-DRUM	20	-0.0113	-	7470A	PPB	26 Oct 2022 10:13:29	U	SMPL
N5253-02 - 1	164718	36	-0.0081	-	7470A	PPB	26 Oct 2022 10:15:46	U	SMPL
N5253-04 - 1	124055	55	-0.0042	-	7470A	PPB	26 Oct 2022 10:18:04	U	SMPL
N5253-06 - 1	124042	49	-0.0054	-	7470A	PPB	26 Oct 2022 10:20:19	U	SMPL
CCV81 - 1	CCV81	24882	5.0314	-	7470A	PPB	26 Oct 2022 10:22:35	U	SMPL
CCB81 - 1	CCB81	-105	-0.0367	-	7470A	PPB	26 Oct 2022 10:24:50	U	SMPL
N5253-08 - 1	124019	44	-0.0064	-	7470A	PPB	26 Oct 2022 10:27:06	U	SMPL
N5254-03 - 1	WC-1-102122	46	-0.0060	-	7470A	PPB	26 Oct 2022 10:29:21	U	SMPL
N5254-06 - 1	WC-2-102122	43	-0.0066	-	7470A	PPB	26 Oct 2022 10:31:37	U	SMPL
N5254-09 - 1	WC-3-102122	162	0.0175	-	7470A	PPB	26 Oct 2022 10:33:52	U	SMPL
N5254-12 - 1	WC-4-102122	71	-0.0010	-	7470A	PPB	26 Oct 2022 10:36:08	U	SMPL
N5254-15 - 1	WC-5-102122	42	-0.0068	-	7470A	PPB	26 Oct 2022 10:38:25	U	SMPL
N5256-02 - 1	SB-F1-COMP	40	-0.0073	-	7470A	PPB	26 Oct 2022 10:40:41	U	SMPL
N5256-04 - 1	SB-F2-COMP	32	-0.0089	-	7470A	PPB	26 Oct 2022 10:42:58	U	SMPL
N5256-06 - 1	SB-H1-COMP	56	-0.0040	-	7470A	PPB	26 Oct 2022 10:45:15	U	SMPL
N5256-08 - 1	SB-N1-COMP	43	-0.0066	-	7470A	PPB	26 Oct 2022 10:47:31	U	SMPL
CCV82 - 1	CCV82	25010	5.0574	-	7470A	PPB	26 Oct 2022 10:49:46	U	SMPL
CCB82 - 1	CCB82	-98	-0.0352	-	7470A	PPB	26 Oct 2022 10:52:02	U	SMPL
N5256-10 - 1	SB-N2-COMP	30	-0.0093	-	7470A	PPB	26 Oct 2022 10:54:17	U	SMPL
N5256-12 - 1	SB-N3-COMP	30	-0.0093	-	7470A	PPB	26 Oct 2022 10:56:33	U	SMPL
N5265-02 - 1	72-11976	-351	-0.0866	-	7470A	PPB	26 Oct 2022 10:58:49	U	SMPL
N5265-02DUP - 1	72-11976DUP	42	-0.0068	-	7470A	PPB	26 Oct 2022 11:01:04	U	SMPL
N5265-02MS - 1	72-11976MS	20358	4.1138	-	7470A	PPB	26 Oct 2022 11:03:20	U	SMPL
N5265-02MSD - 1	72-11976MSD	19623	3.9648	-	7470A	PPB	26 Oct 2022 11:05:36	U	SMPL
PB122528TB - 1	PB122528TB	-36	-0.0227	-	7470A	PPB	26 Oct 2022 11:07:53	U	SMPL
N5265-02LX5 - 1		-166	-0.0490	-	7470A	PPB	26 Oct 2022 11:23:22	U	SMPL
N5265-02A - 1		20230	4.0879	-	7470A	PPB	26 Oct 2022 11:25:41	U	SMPL
CCV83 - 1	CCV83	24993	5.0539	-	7470A	PPB	26 Oct 2022 11:27:56	U	SMPL
CCB83 - 1	CCB83	-58	-0.0271	-	7470A	PPB	26 Oct 2022 11:30:12	U	SMPL