



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122618

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID:

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	10/26/2022	11:00
2	N5207-01		1	13	100.00	>212.0	>212.0	10/26/2022	11:35
3	N5207-01DUP		1	12	100.00	>212.0	>212.0	10/26/2022	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB122618

WorkList Name : FP-102622 WorkList ID : 164389 Department : Wet-Chemistry Date : 10-26-2022 10:57:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/26/2022	Solid	N5207-01	Flash Point	Cool 4 deg C	PSEG03	D11	2689	10/19/2022	1010B

Date/Time 10/26/22 10:50
 Raw Sample Received by: CSN
 Raw Sample Relinquished by: CSN

Date/Time 10/26/22 12:30
 Raw Sample Received by: CSN
 Raw Sample Relinquished by: CSN