



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122622

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP99624
Turbidity Calibration std, 0NTU	WP99623
Turbidity Calibration std, 20NTU	WP99618
Turbidity Calibration std, 40NTU	WP99617
Turbidity Calibration std, 1NTU	WP99621
Turbidity Calibration std, 5NTU	WP99620
Turbidity Calibration - CCV std, 10 NTU	WP99619
10 NTU Standard 500 ml	W2971

Intercept: 0.062

Slope: 0.9299

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.09		10/26/2022	11:10
2	CAL2	1	1	1.130	1.15	-14.9	10/26/2022	11:14
3	CAL3	5	1	4.410	4.68	6.5	10/26/2022	11:18
4	CAL4	10	1	9.460	10.11	-1.1	10/26/2022	11:22
5	CAL5	20	1	18.600	19.94	0.3	10/26/2022	11:26
6	CAL6	40	1	37.300	40.05	-0.1	10/26/2022	11:30



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB122622

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.360	9.999	10/26/2022	11:34
2	ICB		1	0.150	0.095	10/26/2022	11:38
3	CCV1	10	1	9.280	9.913	10/26/2022	11:42
4	CCB1		1	0.152	0.097	10/26/2022	11:46
5	LB122622BL		1	0.150	0.095	10/26/2022	11:50
6	N4955-12		1	0.521	0.494	10/26/2022	11:54
7	CCV2	10	1	9.260	9.891	10/26/2022	12:00
8	CCB2		1	0.154	0.099	10/26/2022	12:04