


LB122624

7471B

INSTRUMENT ID: CVL



Linear

A= 0.0000e+000

B= 1.7616e-004

C= 2.1338e-002

Rho= 0.9999687

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.019	0.019	-13	0.000	-13					
0.2	0.200	0.177	-0.023	882	0.0 %	882					-12
2.5	2.500	2.477	-0.023	13939	0.0 %	13939					-1
5.0	5.000	5.055	0.055	28577	0.0 %	28577					0
7.5	7.500	7.476	-0.024	42316	0.0 %	42316					0
10.0	10.000	9.996	-0.004	56624	0.0 %	56624					0

LB122624 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Oct 2022 15:08:51

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-13	-	0.0000	7471B	PPB	26 Oct 2022 15:16:56	S	Std
0.2 - 1	50.2	882	-	0.2000	7471B	PPB	26 Oct 2022 15:19:14	S	Std
2.5 - 1	52.5	13939	-	2.5000	7471B	PPB	26 Oct 2022 15:21:31	S	Std
5.0 - 1	55	28577	-	5.0000	7471B	PPB	26 Oct 2022 15:23:48	S	Std
7.5 - 1	57.5	42316	-	7.5000	7471B	PPB	26 Oct 2022 15:26:05	S	Std
10.0 - 1	511	56624	-	10.0000	7471B	PPB	26 Oct 2022 15:30:56	S	Std
ICV93 - 1	ICV93	23445	4.1514	-	7471B	PPB	26 Oct 2022 15:33:34	U	SMPL
ICB93 - 1	ICB93	-132	-0.0019	-	7471B	PPB	26 Oct 2022 15:35:49	U	SMPL
CCV84 - 1	CCV84	28532	5.0475	-	7471B	PPB	26 Oct 2022 15:38:06	U	SMPL
CCB84 - 1	CCB84	-115	0.0011	-	7471B	PPB	26 Oct 2022 15:40:21	U	SMPL
CRA - 1	CRA	896	0.1792	-	7471B	PPB	26 Oct 2022 15:42:37	U	SMPL
HighStd - 1	HighStd	56559	9.9848	-	7471B	PPB	26 Oct 2022 15:44:53	U	SMPL
ChkStd - 1	ChkStd	39131	6.9147	-	7471B	PPB	26 Oct 2022 15:47:06	U	SMPL
PB148586BL - 1	PBS	-187	-0.0116	-	7471B	PPB	26 Oct 2022 15:49:26	U	SMPL
PB148586BS - 1	LCSS	22365	3.9612	-	7471B	PPB	26 Oct 2022 15:51:44	U	SMPL
N5218-01 - 1	SP-1	2365	0.4380	-	7471B	PPB	26 Oct 2022 15:54:00	U	SMPL
N5218-01DUP - 1	SP-1DUP	2266	0.4205	-	7471B	PPB	26 Oct 2022 15:56:18	U	SMPL
N5218-01MS - 1	SP-1MS	25884	4.5811	-	7471B	PPB	26 Oct 2022 15:58:35	U	SMPL
N5218-01MSD - 1	SP-1MSD	23934	4.2376	-	7471B	PPB	26 Oct 2022 16:00:52	U	SMPL
N5218-11 - 1	SP-2	2309	0.4281	-	7471B	PPB	26 Oct 2022 16:03:08	U	SMPL
CCV85 - 1	CCV85	28784	5.0919	-	7471B	PPB	26 Oct 2022 16:05:25	U	SMPL
CCB85 - 1	CCB85	-121	0.0000	-	7471B	PPB	26 Oct 2022 16:07:40	U	SMPL
N5218-21 - 1	SP-3	3238	0.5917	-	7471B	PPB	26 Oct 2022 16:09:57	U	SMPL
N5218-31 - 1	SP-4	1240	0.2398	-	7471B	PPB	26 Oct 2022 16:12:13	U	SMPL
N5220-01 - 1	HFI-CON-1020	547	0.1177	-	7471B	PPB	26 Oct 2022 16:14:28	U	SMPL
N5221-01 - 1	CS-1-(8-POINT-COMPOSITE)-PILE-1	7071	1.2670	-	7471B	PPB	26 Oct 2022 16:16:43	U	SMPL
N5221-02 - 1	CS-2-(5-POINT-COMPOSITE)-PILE-2	16148	2.8660	-	7471B	PPB	26 Oct 2022 16:18:59	U	SMPL
N5221-03 - 1	CS-3-(8-POINT-COMPOSITE)-PILE-3	4257	0.7713	-	7471B	PPB	26 Oct 2022 16:21:16	U	SMPL
N5221-04 - 1	CS-4-(8-POINT-COMPOSITE)-PILE-4	8832	1.5772	-	7471B	PPB	26 Oct 2022 16:23:33	U	SMPL
N5238-01 - 1	CHRT-20430	2335	0.4327	-	7471B	PPB	26 Oct 2022 16:25:50	U	SMPL
N5238-03 - 1	2673	8195	1.4650	-	7471B	PPB	26 Oct 2022 16:28:07	U	SMPL
N5238-05 - 1	RBR-251375	1120	0.2186	-	7471B	PPB	26 Oct 2022 16:30:22	U	SMPL
CCV86 - 1	CCV86	29106	5.1487	-	7471B	PPB	26 Oct 2022 16:32:38	U	SMPL
CCB86 - 1	CCB86	-105	0.0028	-	7471B	PPB	26 Oct 2022 16:34:54	U	SMPL
N5218-01LX5 - 1		285	0.0715	-	7471B	PPB	26 Oct 2022 16:37:11	U	SMPL
N5218-01A - 1		23774	4.2094	-	7471B	PPB	26 Oct 2022 16:39:26	U	SMPL
CCV87 - 1	CCV87	29143	5.1552	-	7471B	PPB	26 Oct 2022 16:41:41	U	SMPL
CCB87 - 1	CCB87	-119	0.0004	-	7471B	PPB	26 Oct 2022 16:43:58	U	SMPL