



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB122633

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC pH meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/27/2022	08:35
2	CAL2	1	Water	NA	NA	20.7	7.01	10/27/2022	08:36
3	CAL3	1	Water	NA	NA	20.7	10.03	10/27/2022	08:37
4	ICV	1	Water	NA	NA	20.8	6.99	10/27/2022	08:40
5	CCV1	1	Water	NA	NA	20.7	2.01	10/27/2022	08:42
6	N5282-03	1	Water	NA	NA	22.7	7.72	10/27/2022	08:45
7	N5288-01	1	Water	NA	NA	22.8	7.75	10/27/2022	08:46
8	N5288-02	1	Water	NA	NA	20.3	7.82	10/27/2022	08:50
9	N5290-01	1	Water	NA	NA	20.4	7.88	10/27/2022	08:52
10	N5290-02	1	Water	NA	NA	20.4	7.88	10/27/2022	08:53
11	N5292-01	1	Water	NA	NA	20.2	7.83	10/27/2022	09:00
12	N5292-02	1	Water	NA	NA	20.2	7.56	10/27/2022	09:02
13	N5303-05	1	Water	NA	NA	20.8	6.99	10/27/2022	09:05
14	N5303-06	1	Water	NA	NA	20.9	7.53	10/27/2022	09:10
15	N5303-07	1	Water	NA	NA	21.3	7.17	10/27/2022	09:13
16	CCV2	1	Water	NA	NA	20.9	12.03	10/27/2022	09:15
17	N5309-01	1	Water	NA	NA	20.7	5.11	10/27/2022	09:20
18	N5309-02	1	Water	NA	NA	20.7	5.17	10/27/2022	09:25
19	N5309-03	1	Water	NA	NA	20.8	5.06	10/27/2022	09:30
20	N5309-03DUP	1	Water	NA	NA	20.8	5.07	10/27/2022	09:32

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.8	2.01	10/27/2022	09:35

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n5275

WorkList ID : 164399

Department : Wet-Chemistry

Date : 10-27-2022 08:30:47

VB 122639-38
12/22/2022

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/25/2022	Water	N5282-03	pH	Cool 4 deg C	PSEG03	F11	SUMP-WATER	10/25/2022	9040C
10/25/2022	Water	N5288-01	pH	Cool 4 deg C	ENVI70	F11	252672-7	10/25/2022	9040C
10/25/2022	Water	N5288-02	pH	Cool 4 deg C	ENVI70	F11	252672-8	10/25/2022	9040C
10/25/2022	Water	N5290-01	pH	Cool 4 deg C	ENVI70	F11	257248-7	10/25/2022	9040C
10/25/2022	Water	N5290-02	pH	Cool 4 deg C	ENVI70	F11	257248-8	10/25/2022	9040C
10/25/2022	Water	N5292-01	pH	Cool 4 deg C	ENVI70	F11	251032-3	10/25/2022	9040C
10/25/2022	Water	N5292-02	pH	Cool 4 deg C	ENVI70	F11	251032-4	10/25/2022	9040C
10/26/2022	Water	N5303-05	pH	Cool 4 deg C	PSEG03	F11	NWB-1837	10/26/2022	9040C
10/26/2022	Water	N5303-06	pH	Cool 4 deg C	PSEG03	F11	NWB-1838	10/26/2022	9040C
10/26/2022	Water	N5303-07	pH	Cool 4 deg C	PSEG03	F11	NWB-1842	10/26/2022	9040C
10/26/2022	Water	N5309-01	pH	Cool 4 deg C	PSEG03	F11	34247	10/26/2022	9040C
10/26/2022	Water	N5309-02	pH	Cool 4 deg C	PSEG03	F11	34248	10/26/2022	9040C
10/26/2022	Water	N5309-03	pH	Cool 4 deg C	PSEG03	F11	34249	10/26/2022	9040C

Date/Time 10/27/22 08:32

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 10/27/22 11:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]