


LB122636

7471B

INSTRUMENT ID: CVL

Linear



A= 0.0000e+000

B= 1.7365e-004

C= 1.3357e-002

Rho= 0.9996051

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.003	0.003	-59	0.000	-59					-
0.2	0.200	0.175	-0.025	932	0.0 %	932					-13
2.5	2.500	2.468	-0.032	14136	0.0 %	14136					-1
5.0	5.000	5.186	0.186	29786	0.0 %	29786					4
7.5	7.500	7.335	-0.165	42162	0.0 %	42162					-2
10.0	10.000	10.033	0.033	57700	0.0 %	57700					0

LB122636 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Oct 2022 10:52:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	-59	-	0.0000	7471B	PPB	27 Oct 2022 11:06:53	S	Std
0.2 - 1	58.2	932	-	0.2000	7471B	PPB	27 Oct 2022 11:09:10	S	Std
2.5 - 1	52.5	14136	-	2.5000	7471B	PPB	27 Oct 2022 11:11:27	S	Std
5.0 - 1	55	29786	-	5.0000	7471B	PPB	27 Oct 2022 11:13:44	S	Std
7.5 - 1	59.5	42162	-	7.5000	7471B	PPB	27 Oct 2022 11:16:01	S	Std
10.0 - 1	314	57700	-	10.0000	7471B	PPB	27 Oct 2022 11:18:17	S	Std
ICV94 - 1	ICV94	24083	4.1954	-	7471B	PPB	27 Oct 2022 11:21:05	U	SMPL
ICB94 - 1	ICB94	-130	-0.0092	-	7471B	PPB	27 Oct 2022 11:23:20	U	SMPL
CCV88 - 1	CCV88	29892	5.2041	-	7471B	PPB	27 Oct 2022 11:25:35	U	SMPL
CCB88 - 1	CCB88	-125	-0.0083	-	7471B	PPB	27 Oct 2022 11:27:50	U	SMPL
CRA - 1	CRA	911	0.1716	-	7471B	PPB	27 Oct 2022 11:30:06	U	SMPL
HighStd - 1	HighStd	57924	10.0719	-	7471B	PPB	27 Oct 2022 11:32:21	U	SMPL
ChkStd - 1	ChkStd	39602	6.8903	-	7471B	PPB	27 Oct 2022 11:34:37	U	SMPL
PB148587BL - 1	PBS	-181	-0.0181	-	7471B	PPB	27 Oct 2022 11:36:55	U	SMPL
PB148587BS - 1	LCSS	23036	4.0136	-	7471B	PPB	27 Oct 2022 11:39:13	U	SMPL
N5263-01 - 1	31502	567	0.1118	-	7471B	PPB	27 Oct 2022 11:41:29	U	SMPL
N5265-01 - 1	72-11976	4588	0.8101	-	7471B	PPB	27 Oct 2022 11:43:46	U	SMPL
N5265-01DUP - 1	72-11976DUP	4652	0.8212	-	7471B	PPB	27 Oct 2022 11:46:03	U	SMPL
N5265-01MS - 1	72-11976MS	28482	4.9593	-	7471B	PPB	27 Oct 2022 11:48:20	U	SMPL
N5265-01MSD - 1	72-11976MSD	28307	4.9289	-	7471B	PPB	27 Oct 2022 11:50:35	U	SMPL
CCV89 - 1	CCV89	30108	5.2416	-	7471B	PPB	27 Oct 2022 11:52:53	U	SMPL
CCB89 - 1	CCB89	9	0.0149	-	7471B	PPB	27 Oct 2022 11:57:19	U	SMPL
N5265-03 - 1	CONCRETE-72-11976	918	0.1728	-	7471B	PPB	27 Oct 2022 11:59:34	U	SMPL
N5267-01 - 1	TP1-1024	10118	1.7704	-	7471B	PPB	27 Oct 2022 12:01:49	U	SMPL
N5267-04 - 1	TP2-1024	21586	3.7618	-	7471B	PPB	27 Oct 2022 12:04:04	U	SMPL
N5267-07 - 1	TP3-1024	3446	0.6118	-	7471B	PPB	27 Oct 2022 12:06:20	U	SMPL
N5267-10 - 1	TP4-1024	23048	4.0157	-	7471B	PPB	27 Oct 2022 12:08:36	U	SMPL
N5267-13 - 1	TP5-1024	2468	0.4419	-	7471B	PPB	27 Oct 2022 12:10:53	U	SMPL
N5267-16 - 1	TP6-1024	1797	0.3254	-	7471B	PPB	27 Oct 2022 12:13:09	U	SMPL
N5267-19 - 1	TP7-1024	2215	0.3980	-	7471B	PPB	27 Oct 2022 12:15:26	U	SMPL
N5268-01 - 1	OK-03-102422	2334	0.4187	-	7471B	PPB	27 Oct 2022 12:17:43	U	SMPL
N5269-01 - 1	SU-03-102422	2119	0.3813	-	7471B	PPB	27 Oct 2022 12:19:58	U	SMPL
CCV90 - 1	CCV90	30020	5.2264	-	7471B	PPB	27 Oct 2022 12:22:14	U	SMPL
CCB90 - 1	CCB90	-17	0.0104	-	7471B	PPB	27 Oct 2022 12:26:42	U	SMPL
N5282-01 - 1	SUMP-STONE	49	0.0219	-	7471B	PPB	27 Oct 2022 12:28:57	U	SMPL
N5283-01 - 1	IRV-SOIL-1025	977	0.1830	-	7471B	PPB	27 Oct 2022 12:31:12	U	SMPL
N5284-01 - 1	MH-1-WC	273	0.0608	-	7471B	PPB	27 Oct 2022 12:33:27	U	SMPL
N5284-04 - 1	MH-3-WC	1030	0.1922	-	7471B	PPB	27 Oct 2022 12:35:43	U	SMPL
N5286-01 - 1	BBS-SOILCUT-1026	203612	35.3708	-	7471B	PPB	27 Oct 2022 12:37:59	U	SMPL
N5308-01 - 1	PL-2-102622	1319	0.2424	-	7471B	PPB	27 Oct 2022 12:40:15	U	SMPL
N5265-01LX5 - 1		683	0.1320	-	7471B	PPB	27 Oct 2022 12:42:36	U	SMPL
N5265-01A - 1		28044	4.8832	-	7471B	PPB	27 Oct 2022 12:44:52	U	SMPL
N5286-01DLX10 - 1		21255	3.7043	-	7471B	PPB	27 Oct 2022 12:47:09	U	SMPL
CCV91 - 1	CCV91	30115	5.2429	-	7471B	PPB	27 Oct 2022 12:49:26	U	SMPL
CCB91 - 1	CCB91	-248	-0.0297	-	7471B	PPB	27 Oct 2022 12:51:43	U	SMPL