



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 10/27/2022 16:25 TEMP1 OUT: 103 °C 10/28/2022 08:00
TEMP2 IN: 103 °C 10/28/2022 08:30 TEMP2 OUT: 103 °C 10/28/2022 10:00
TEMP3 IN: 540 °C 10/28/2022 10:30 TEMP3 OUT: 540 °C 10/28/2022 12:00
TEMP4 IN: 560 °C 10/28/2022 12:30 TEMP4 OUT: 560 °C 10/28/2022 14:00

Run Number: LB122643

SUPERVISOR: Iwona

ANALYST: JIGNESH

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB122643BL	89.6753	89.6753	100	89.6753	89.6753	89.6753	89.6753	0.0000	0
2	N5304-01	68.7763	68.7763	100	68.9951	68.9951	68.9633	68.9633	0.0318	318
3	N5304-01DUP	91.5299	91.5299	100	91.7522	91.7522	91.7197	91.7197	0.0325	325

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

18722643

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tvs n5304 w WorkList ID : 164418 Department : Wet-Chemistry Date : 10-27-2022 13:06:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/02/2022	Water	N5304-01	TVS	Cool 4 deg C	PSEG03	F11	DEEP-WELL	10/26/2022	SM2540 G

Date/Time 10/27/22 13:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 10/27/22 16:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]