

L9122647


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

45677



A= 0.0000e+000

B= 2.1583e-004

C= -1.5060e-002

Rho= 0.9996065

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.015	-0.015	-2	0.000	-2					
0.2	0.200	0.183	-0.017	917	0.0 %	917					19
2.5	2.500	2.432	-0.068	11339	0.0 %	11339					13
5.0	5.000	5.101	0.101	23705	0.0 %	23705					22
7.5	7.500	7.656	0.156	35543	0.0 %	35543					22
10.0	10.000	9.843	-0.157	45677	0.0 %	45677					12

LB122647 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Oct 2022 08:07:16

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-2	-	0.00007470A	PPB	28 Oct 2022 08:25:32	S	Std
0.2 - 1	50.2	917	-	0.20007470A	PPB	28 Oct 2022 08:27:50	S	Std
2.5 - 1	50.5	11339	-	2.50007470A	PPB	28 Oct 2022 08:30:07	S	Std
5.0 - 1	55	23705	-	5.00007470A	PPB	28 Oct 2022 08:32:24	S	Std
7.5 - 1	57.5	35543	-	7.50007470A	PPB	28 Oct 2022 08:34:41	S	Std
10.0 - 1	58	45677	-	10.00007470A	PPB	28 Oct 2022 08:36:56	S	Std
ICV95 - 1	ICV95	18242	3.9221	-7470A	PPB	28 Oct 2022 08:41:53	U	SMPL
ICB95 - 1	ICB95	-71	-0.0304	-7470A	PPB	28 Oct 2022 08:44:07	U	SMPL
CCV92 - 1	CCV92	23573	5.0726	-7470A	PPB	28 Oct 2022 08:46:22	U	SMPL
CCB92 - 1	CCB92	-83	-0.0330	-7470A	PPB	28 Oct 2022 08:48:37	U	SMPL
CRA - 1	CRA	929	0.1854	-7470A	PPB	28 Oct 2022 08:50:53	U	SMPL
HighStd - 1	HighStd	45731	9.8549	-7470A	PPB	28 Oct 2022 08:53:08	U	SMPL
ChkStd - 1	ChkStd	33001	7.1074	-7470A	PPB	28 Oct 2022 08:55:23	U	SMPL
PB148530BL - 1	PBW	-147	-0.0468	-7470A	PPB	28 Oct 2022 08:57:40	U	SMPL
PB148530BS - 1	LCSW	17664	3.7973	-7470A	PPB	28 Oct 2022 08:59:56	U	SMPL
N5185-02 - 1	MW-01	-144	-0.0461	-7470A	PPB	28 Oct 2022 09:02:12	U	SMPL
N5185-04 - 1	MW-02	-83	-0.0330	-7470A	PPB	28 Oct 2022 09:04:29	U	SMPL
N5185-06 - 1	MW-04	-103	-0.0373	-7470A	PPB	28 Oct 2022 09:06:46	U	SMPL
N5185-08 - 1	MW-05	-81	-0.0325	-7470A	PPB	28 Oct 2022 09:09:03	U	SMPL
N5185-10 - 1	MW-08	-89	-0.0343	-7470A	PPB	28 Oct 2022 09:11:18	U	SMPL
CCV93 - 1	CCV93	23466	5.0495	-7470A	PPB	28 Oct 2022 09:13:33	U	SMPL
CCB93 - 1	CCB93	-66	-0.0293	-7470A	PPB	28 Oct 2022 09:15:49	U	SMPL
N5185-12 - 1	MW-10	31	-0.0084	-7470A	PPB	28 Oct 2022 09:18:04	U	SMPL
N5185-12DUP - 1	MW-10DUP	22	-0.0103	-7470A	PPB	28 Oct 2022 09:20:20	U	SMPL
N5185-15 - 1	N5185-12MS	18279	3.9300	-7470A	PPB	28 Oct 2022 09:22:35	U	SMPL
N5185-16 - 1	N5185-12MSD	17842	3.8357	-7470A	PPB	28 Oct 2022 09:24:51	U	SMPL
N5185-18 - 1	PZ-02	-158	-0.0492	-7470A	PPB	28 Oct 2022 09:27:07	U	SMPL
N5185-20 - 1	PZ-3R	16	-0.0116	-7470A	PPB	28 Oct 2022 09:29:23	U	SMPL
N5185-22 - 1	RW-01	-102	-0.0371	-7470A	PPB	28 Oct 2022 09:31:40	U	SMPL
N5185-24 - 1	RW-02	-98	-0.0362	-7470A	PPB	28 Oct 2022 09:33:57	U	SMPL
N5282-03 - 1	SUMP-WATER	25	-0.0097	-7470A	PPB	28 Oct 2022 09:36:14	U	SMPL
N5288-01 - 1	252672-7	-89	-0.0343	-7470A	PPB	28 Oct 2022 09:38:29	U	SMPL
CCV94 - 1	CCV94	23153	4.9820	-7470A	PPB	28 Oct 2022 09:40:44	U	SMPL
CCB94 - 1	CCB94	-100	-0.0366	-7470A	PPB	28 Oct 2022 09:43:00	U	SMPL
N5288-02 - 1	252672-8	-91	-0.0347	-7470A	PPB	28 Oct 2022 09:45:16	U	SMPL
N5290-01 - 1	257248-7	-91	-0.0347	-7470A	PPB	28 Oct 2022 09:47:32	U	SMPL
N5290-02 - 1	257248-8	-162	-0.0500	-7470A	PPB	28 Oct 2022 09:49:47	U	SMPL
N5292-01 - 1	251032-3	-84	-0.0332	-7470A	PPB	28 Oct 2022 09:52:03	U	SMPL
N5292-02 - 1	251032-4	-64	-0.0289	-7470A	PPB	28 Oct 2022 09:54:19	U	SMPL
PB148594BL - 1	PBW	-16	-0.0185	-7470A	PPB	28 Oct 2022 09:56:36	U	SMPL
PB148594BS - 1	LCSW	17637	3.7915	-7470A	PPB	28 Oct 2022 09:58:52	U	SMPL
N5267-02 - 1	TP1-1024	-104	-0.0375	-7470A	PPB	28 Oct 2022 10:01:09	U	SMPL
N5267-05 - 1	TP2-1024	-8	-0.0168	-7470A	PPB	28 Oct 2022 10:03:26	U	SMPL
N5267-08 - 1	TP3-1024	-21	-0.0196	-7470A	PPB	28 Oct 2022 10:05:41	U	SMPL
CCV95 - 1	CCV95	23769	5.1149	-7470A	PPB	28 Oct 2022 10:07:57	U	SMPL
CCB95 - 1	CCB95	-109	-0.0386	-7470A	PPB	28 Oct 2022 10:10:12	U	SMPL
N5267-11 - 1	TP4-1024	-47	-0.0252	-7470A	PPB	28 Oct 2022 10:12:28	U	SMPL
N5267-14 - 1	TP5-1024	-65	-0.0291	-7470A	PPB	28 Oct 2022 10:14:43	U	SMPL
N5267-17 - 1	TP6-1024	-31	-0.0218	-7470A	PPB	28 Oct 2022 10:16:59	U	SMPL
N5267-20 - 1	TP7-1024	-38	-0.0233	-7470A	PPB	28 Oct 2022 10:19:14	U	SMPL
N5268-04 - 1	OK-03-102422	-74	-0.0310	-7470A	PPB	28 Oct 2022 10:21:30	U	SMPL
N5271-01 - 1	WC-IDW-0001	-36	-0.0228	-7470A	PPB	28 Oct 2022 10:23:46	U	SMPL
N5273-02 - 1	LIDEN-GAS-PIPELINE	-51	-0.0261	-7470A	PPB	28 Oct 2022 10:26:02	U	SMPL
N5284-02 - 1	MH-1-WC	-44	-0.0246	-7470A	PPB	28 Oct 2022 10:28:19	U	SMPL
N5284-05 - 1	MH-3-WC	-46	-0.0250	-7470A	PPB	28 Oct 2022 10:30:36	U	SMPL
N5284-05DUP - 1	MH-3-WCDUP	-36	-0.0228	-7470A	PPB	28 Oct 2022 10:32:52	U	SMPL
CCV96 - 1	CCV96	24434	5.2585	-7470A	PPB	28 Oct 2022 10:35:07	U	SMPL
CCB96 - 1	CCB96	-90	-0.0345	-7470A	PPB	28 Oct 2022 10:37:23	U	SMPL
N5284-05MS - 1	MH-3-WCMS	21944	4.7210	-7470A	PPB	28 Oct 2022 10:39:38	U	SMPL
N5284-05MSD - 1	MH-3-WCMSD	19573	4.2093	-7470A	PPB	28 Oct 2022 10:41:54	U	SMPL
PB148572TB - 1	PB148572TB	-112	-0.0392	-7470A	PPB	28 Oct 2022 10:44:09	U	SMPL
N5185-12LX5 - 1		-133	-0.0438	-7470A	PPB	28 Oct 2022 10:46:25	U	SMPL
N5185-12A - 1		18408	3.9579	-7470A	PPB	28 Oct 2022 10:48:40	U	SMPL
N5284-05LX5 - 1		-185	-0.0550	-7470A	PPB	28 Oct 2022 10:50:57	U	SMPL
N5284-05A - 1		18811	4.0449	-7470A	PPB	28 Oct 2022 10:53:13	U	SMPL
CCV97 - 1	CCV97	24290	5.2274	-7470A	PPB	28 Oct 2022 10:55:30	U	SMPL
CCB97 - 1	CCB97	-103	-0.0373	-7470A	PPB	28 Oct 2022 10:57:47	U	SMPL