



## Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122648

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

| Reagent/Standard | Lot/Log # |
|------------------|-----------|
| p-xylene (ICV)   | W2744     |

| Seq | LabID       | True Value °F | DL | Initial Sample °C | Celsius °C | Result °F | Final Result °F | Anal Date  | Anal Time |
|-----|-------------|---------------|----|-------------------|------------|-----------|-----------------|------------|-----------|
| 1   | ICV         | 81            | 1  | 8                 | 28.00      | 82.4      | 82.4            | 10/27/2022 | 15:30     |
| 2   | N5303-04    |               | 1  | 14                | 62.00      | 143.6     | 143.6           | 10/27/2022 | 16:05     |
| 3   | N5303-04DUP |               | 1  | 14                | 62.00      | 143.6     | 143.6           | 10/27/2022 | 16:30     |

Result = (Celsius \* 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) \* 0.06

LB122648

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FLASH-1027      WorkList ID : 164424      Department : Wet-Chemistry      Date : 10-27-2022 15:59:06

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|-----------------------------|-----------------|--------------|--------|
| 11/02/2022 | Solid  | N5303-04 | Flash Point | Cool 4 deg C | PSEG03   | F11                         | NWB-1821        | 10/26/2022   | 1010B  |

Date/Time 10/27/22 16100  
Raw Sample Received by: SL (WC)  
Raw Sample Relinquished by: CP (SM)

Date/Time 10/27/2022 17:30  
Raw Sample Received by: CP (SM)  
Raw Sample Relinquished by: SL (WC)