



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: jignesh

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB122649

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	10/28/2022	09:30
2	N5282-03		1	12	100.00	>212.0	>212.0	10/28/2022	09:50
3	N5303-05		1	14	100.00	>212.0	>212.0	10/28/2022	10:20
4	N5303-06		1	13	100.00	>212.0	>212.0	10/28/2022	10:50
5	N5303-07		1	12	100.00	>212.0	>212.0	10/28/2022	11:20
6	N5304-01		1	13	100.00	>212.0	>212.0	10/28/2022	11:50
7	N5309-01		1	15	100.00	>212.0	>212.0	10/28/2022	12:20
8	N5309-02		1	13	100.00	>212.0	>212.0	10/28/2022	13:05
9	N5309-03		1	14	100.00	>212.0	>212.0	10/28/2022	13:35
10	N5329-01		1	14	100.00	>212.0	>212.0	10/28/2022	14:05
11	N5329-03		1	12	100.00	>212.0	>212.0	10/28/2022	14:35
12	N5338-01		1	14	100.00	>212.0	>212.0	10/28/2022	15:05
13	N5338-01DUP		1	13	100.00	>212.0	>212.0	10/28/2022	15:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06