



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB122651

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2905
BUFFER PH 7.00 GREEN 1PINT PK6	W2906
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	10/27/2022	16:25
2	CAL2	1	Water	NA	NA	20.7	7.00	10/27/2022	16:26
3	CAL3	1	Water	NA	NA	20.7	10.03	10/27/2022	16:27
4	ICV	1	Water	NA	NA	20.7	6.99	10/27/2022	16:30
5	CCV1	1	Water	NA	NA	20.7	2.01	10/27/2022	16:32
6	N5313-02	1	Solid	20.02	20	22.6	8.46	10/27/2022	16:35
7	N5313-09	1	Solid	20.01	20	22.5	8.56	10/27/2022	16:36
8	N5313-16	1	Solid	20.03	20	22.5	8.48	10/27/2022	16:39
9	N5316-02	1	Solid	20.05	20	21.8	8.52	10/27/2022	16:40
10	N5316-09	1	Solid	20.03	20	22.3	8.65	10/27/2022	16:42
11	N5332-02	1	Solid	20.02	20	22.2	6.56	10/27/2022	16:45
12	N5332-02DUP	1	Solid	20.02	20	22.3	6.57	10/27/2022	16:46
13	CCV2	1	Water	NA	NA	20.8	12.03	10/27/2022	16:49

W122651
W122651

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N5332 WorkList ID : 164426 Department : Wet-Chemistry Date : 10-27-2022 15:59:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/26/2022	Solid	N5313-02	Corrosivity	Cool 4 deg C	SOUN03	F11	COMPOSITE-SW-QUARTER	10/26/2022	9045D
10/26/2022	Solid	N5313-09	Corrosivity	Cool 4 deg C	SOUN03	F11	COMPOSITE-NW-QUARTER	10/26/2022	9045D
10/26/2022	Solid	N5313-16	Corrosivity	Cool 4 deg C	SOUN03	F11	COMPOSITE-SE-QUARTER	10/26/2022	9045D
10/26/2022	Solid	N5316-02	Corrosivity	Cool 4 deg C	SOUN03	F11	COMPOSITE-NE-QUARTER	10/26/2022	9045D
10/26/2022	Solid	N5316-09	Corrosivity	Cool 4 deg C	SOUN03	F11	COMPOSITE-PILE-2	10/26/2022	9045D
10/27/2022	Solid	N5332-02	Corrosivity	Cool 4 deg C	PSEG03	F11	MH-4-WC	10/27/2022	9045D

Date/Time 10/27/22 16:00
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC

Date/Time 10/27/22 18:30
Raw Sample Received by: JDC
Raw Sample Relinquished by: JDC